

FDA- Konformitätserklärung

Produkt

Product

Wir als Hersteller von Produkten für die Analyse erklären, dass die prozessmedienberührenden Komponenten des oben genannten Produktes in ihrer Zusammensetzung der unten aufgeführten Bestimmung der

„Food and Drug Administration (FDA)“

entsprechen.

Diese Herstellererklärung ist nur gültig für das oben genannte Produkt.

Komponente
Component

O-Ring

O-ring

Material
Material

EPDM

E7502

FDA Declaration of Conformity

71307106 Kit CLS82D: Dichtungen EPDM, FDA

71307106 Kit CLS82D: Gaskets EPDM, FDA

We as product manufacturer of analysis products declare that the following type of product consists wetted parts with process contact, which in their composition are in compliance with below mentioned regulation

Food and Drug Administration (FDA).

This manufacturer's declaration is only valid for above mentioned product.

FDA-Konformität
FDA conformity
21 CFR 177.2600

Endress+Hauser Conducta GmbH+Co. KG
Dieselstraße 24, 70839 Gerlingen, Germany

Datum

Date

09.01.2020



i.V. Uwe Röbiger
Manager Certifications & Approvals



i.V. Dr. Julia Groß
Specialist R&D Chemical Sensors

Zertifikats-Nr. | Certificate-no.
FDA_00237_01.20

Endress+Hauser 
People for Process Automation