

Prüf- und Zertifizierungsstelle

ZELM Ex

(1) EC-Type Examination Certificate

(2) Equipment and Protected Systems Intended for Use in Potentially Explosive Atmospheres -
Directive 94/9/EC

(3) EC-Type Examination Certificate Number

ZELM 04 ATEX 0200

(4) Unit **Explosion-protected PDA Type i.roc 6* -Ex / Type i.roc 6*N -Ex**

(5) Manufacturer: **ecom Instruments GmbH**

(6) Address: **D-97959 Assamstadt**

(7) The design of this equipment and its various approved embodiments are defined in the attachment to this type examination certificate.

(8) The Prüf- und Zertifizierungsstelle ZELM, notified body no. 0820 in accordance with Article 9 of the EC Council Directive dated March 23, 1994 (94/9/EC), certifies that this equipment has been found to conform with the essential health and safety requirements for the design and construction of equipment and protected systems for proper intended use in potentially explosive areas in accordance with Appendix II of the directive.

The results of the test are documented in the confidential test report no. ZELM Ex 0160412277.

(9) The essential health and safety requirements are met by virtue of conformity with

EN 50 014: 1997+A1+A2

EN 50 020: 2002

EN 50 281-1-1:1998

(10) If the certification number is followed by an "X", then this indicates that special conditions exist for the safe operation of the equipment. These special conditions are contained in the attachment to this certificate.

(11) This EC-type examination certificate only refers to the design, checking and testing of the specified unit or protected system in accordance with directive 94/9/EC. Further requirements contained in this directive may apply with regard to the manufacturing process and the supply of the equipment or protected system. Such requirements are not covered by this certification.

(12) The equipment must be labelled with the following information:

 **II 2 G EEx ia IIC T4**
II 2 D T 95°C IP65

Zertifizierungsstelle **ZELM Ex**

Signature

Dipl.-Ing. Harald Zelm

Page _

EC-type examination certificates without signature and stamp are not valid.

This EC-Type examination certificate may only be circulated without alteration. Extracts or alterations are subject to approval by the Prüf- und Zertifizierungsstelle ZELM Ex

Prüf- und Zertifizierungsstelle ZELM Ex • Siekgraben 56 • D-38124 Braunschweig Germany

Prüf- und Zertifizierungsstelle

ZELM Ex

Attachment

(13)

(14)

EC-Type Examination Certificate ZELM 04 ATEX 0200

(15) Description of the unit

The explosion-protected PDA (Personal Digital Assistant) is used for mobile data management in potentially explosive atmospheres. Other accessories such as additional memory can be integrated into the unit. The asterisk (*) in the type designation is replaced with a two-digit number. This number is used to specify unit models that do not represent a safety-relevant variation of the device. Data communication with other devices is performed exclusively by means of a wireless connection via integrated Bluetooth / WLAN and IrDa interfaces.

In potentially explosive atmospheres, the approved unit may be operated for the purpose of exchanging data only in conjunction with other devices.

The unit is fitted with a power supply in the form of an integral rechargeable battery. This rechargeable battery is not designed to be replaced.

In potentially explosive atmospheres, the explosion-protected PDA Type i.roc 6*N -Ex may only be cleaned using a damp cloth. The housing of the explosion-protected PDA Type i.roc 6* -Ex is electrostatically conductive. Consequently, the usually required warnings are not necessary in this instance.

The permitted ambient temperature is -10 °C to +50 °C.

Electrical Data

Power supply circuit

from an internal rechargeable battery, EEx ia IIC The internal circuits are intrinsically safe.

WLAN output

Tx power: 15 dBm (31.6 mW)

Bluetooth output

Tx power: 1 mW (Class II)

Note:

The operating instructions must be followed. In particular, the following instructions are to be adhered to:

- The unit must not be opened within the potentially explosive area.
- Charging must be performed outside the potentially explosive area using the designated charger Type LG x10.
- In potentially explosive atmospheres, the explosion-protected PDA Type i.roc 6*N -Ex may only be cleaned using a damp cloth. All instructions printed on the device must be adhered to.

(16) Test report no.

ZELM Ex 0160412277

(17) Special conditions

Not applicable

(18) Essential health and safety requirements

Satisfied by compliance with the relevant standards

Zertifizierungsstelle **ZELM Ex**

Braunschweig, 29.03.2004

Signature

Dipl.-Ing. Harald Zelm

Page 2/2

Prüf- und Zertifizierungsstelle

ZELM Ex

1. Addition

(Addition in accordance with EC Directive 94/9 Appendix III Item 6)

for EC-Type Examination Certificate

ZELM 04 ATEX 0200

Equipment: **Explosion-protected PDA type i.roc 6*-Ex and/or type i.roc 6*N- Ex**
Manufacturer: **ecom Instruments GmbH**
Address: **D-97959 Assamstadt**

Description of the addition

The first addition to the EC-type examination certificate relates to the internal circuit layout and the designation.

The technical data and the information provided on the EC-type examination certificate remain unchanged.

The explosion-protected PDA type i.roc 6* -Ex and/or type i.roc 6*N -Ex may now be manufactured subject to compliance with this modification.

The equipment must now be labelled with the following information:



II 2 G EEx Ia IIC T4
II 2 D T 99°C IP65

Test report no.

ZELM Ex 1080412323

Special conditions

Not applicable

Essential health and safety requirements

The essential health and safety requirements are met by virtue of conformity with the standards listed in the EC-type examination certificate.

Zertifizierungsstelle **ZELM Ex**

Braunschweig, October 25, 2003

Dipl.-Ing. Harald Zelm

Page 1 of 1

Prüf- und Zertifizierungsstelle
ZELM Ex

2nd addition

(Addition in accordance with EC Directive 94/9 Appendix III Item 6)

to EC-Type Examination Certificate

ZELM 04 ATEX 0200

Equipment: Explosion-protected PDA type i.roc 6*-Ex and/or type i.roc 6*N-Ex
Manufacturer: ecom Instruments GmbH
Address: D-97959 Assamstadt

Description of the addition

This 2nd addition to the EC-type examination certificate relates to the integration of an LED scanner.

The technical data and designation as well as the information provided on the EC-type examination certificate remain unchanged.

The explosion-protected PDA type i.roc 6* -Ex and/or type i.roc 6*N -Ex may now be manufactured subject to compliance with this modification.

Test report no. ZELM Ex
1110412328

Special conditions

Not applicable

Essential health and safety requirements

The essential health and safety requirements are met by virtue of conformity with the standards listed in the EC-type examination certificate.

Zertifizierungsstelle ZELM Ex

Braunschweig, 15.12.2004

Prüf- und Zertifizierungsstelle

ZELM Ex

3rd Addition

(Addition in accordance with EC Directive 94/9 Appendix III Item 6)

to the EC-Type Examination

Certificate ZELM 04 ATEX 0200

Equipment: **Explosion-protected PDA type i.roc 6*-Ex and/or i.roc 6*N-Ex**
Manufacturer: **ecom instruments GmbH**
Address: **Industriestr. 2 D-97959 Assamstadt**

Description of the addition

The third addition to the EC-type examination certificate relates specifically to the internal layout without significant changes to the printed circuit board layout. In addition, the equipment may be alternatively fitted with an additional RFID reader and a pistol grip (with button) or with a handle (corresponds to pistol grip, however, without button).

The ignition protection type/designation is as follows:

II 2G	EEx ia	II C T4
II 2D	T99°C	IP 65

The technical data and designation as well as the information provided on the EC-type examination certificate remain unchanged.

Test report no. ZELM Ex

0970512407

Special conditions

Not applicable

Essential health and safety requirements

The essential health and safety requirements are met by virtue of conformity with the standards listed in the EC-type examination certificate.

ZELM certifying body

Braunschweig, July 15,
2005

Prüf- und Zertifizierungsstelle

ZELM Ex

Fourth Addition

(Addition in accordance with EC Directive 94/9 Appendix III Item 6)

to the EC-Type Examination

Certificate ZELM 04 ATEX 0200

Equipment: **Explosion-protected PDA type i.roc 6*-Ex or i.roc 6*N-Ex**
Manufacturer: **ecom instruments GmbH**
Address: **Industriestr. 2, D-97959 Assamstadt**

Description of the addition

The fourth addition to the EC-type examination certificate relates to expanding the explosion-protected PDA by the type i.roc 62. -Ex, which differs from the previous model particularly in its internal layout. This device can be equipped with the previously approved accessories and also, as an alternative, with an additional HART modem.

These possible, approved accessories are listed in the safety instructions.

The designation/ignition protection type for this version is as follows:

II (1)2 G EEx ia IIC T4
II (1)2 D T99°C IP 65

This version of the explosion-protected PDA type i.roc 62. -Ex may also be used with an additional HART modem as an alternative. Measurements on intrinsically safe HART circuits conducted in areas that require Category 1 are approved subject to compliance with the following connection values.

Electrical data:

HART modem

In ignition protection type: intrinsically safe, EEx ia IIC

Maximum values:

$U_o = 1.4 \text{ V}$
 $P_o = 0 \text{ W}$
 $I_o = 0 \text{ mA}$

maximum permissible external capacitance

$C_o = 1000 \text{ }\mu\text{F}$

maximum permissible external inductance

$L_o = 1000 \text{ mH}$

or

for the connection of certified intrinsically safe circuits with the following maximum values:

$U_i = 30 \text{ V}$
 $I_i = 333.3 \text{ mA}$
 $P_i = 1 \text{ W}$

The effective internal inductance L_i and capacitance C_i are negligibly small or measurements are not on intrinsically safe circuits

Measurement inputs

Maximum values

$U_m = 30 \text{ VDC}$
 $I = 333.3 \text{ mA}$

All other electrical data and the type of designation as well as the information provided on the EC-type examination certificate remain unchanged.

Prüf- und Zertifizierungsstelle

ZELM Ex

Fourth Addition to the EC-type examination certificate ZELM 04 ATEX0200

Test report no.

ZELM Ex 0340612459

Special conditions

Not applicable

Essential health and safety requirements

The essential health and safety requirements are met by virtue of conformity with the standards listed in the EC-type examination certificate.

Zertifizierungsstelle **ZELM** ex

Braunschweig, July 12, 2006

Translation