

















Manufacturer Information

for users regarding software updates (following the NAMUR recommendation 53)

\boxtimes	Field	de	vice /	′ signal	process	sing (device	3
				,			/ 4	

☐ Monitoring-/operation system/hand held terminal etc.

☐ Modem / interface

Manufacturer : Endress+Hauser Flowtec AG

Device : Thermal Mass Flow Measuring System

Type : t-mass 65 HART (0x65)

Order code : 65***_******A, B, R, S, T, U, 2, 5, 6, 8

2. Software

Previous firmware version : 1.00.00 New firmware version : 1.00.01

How can the previous : • Local display

firmware version be

• Nameplate for connections

identified? • Configuration and service software e.g. FieldCare

• Handheld terminal for remote communication e.g. Field Xpert

SFX100

Description of the modifications in comparison to the previous

: • Changes to handling of process error message #435

 Changes to software limits in the SENSOR DATA and PROCESS PARAMETER groups

Fieldcheck interface compatibility

3. Compatibility

Is the operating tool compatible with the installed device firmware?

□ Ye

No, description:

firmware version

An update of the DD's (device description) is required for all operating tools in use.



















Manufacturer Information

for users regarding software updates (following the NAMUR recommendation 53)

	Is a firmware update generally recommended?									
		Yes:								
No, reason: If the additional properties are not required for the operation of the measurement system, then no firmware update is necessary.										
								4.	Ins	truction manua
	Is a new instruction manual necessary due to the modification of the firmware?									
	□ Yes									
		No								
	Which manual corresponds to the new firmware?									
	De	vice	Communication option	s Manual	Identification					
	t-m	nass 65	HART	Operating instructions	BA111D/06/en/12.05					
	al									
5.	Pri	ce								
	Ch									

