



Manufacturer Information

for users regarding software updates
(following the NAMUR recommendation 53)

1 Type of device

- ☒ Field device / signal processing device
- ☐ Monitoring- / operation system / hand held terminal etc.
- ☐ Modem / interface

Manufacturer : Endress+Hauser GmbH+Co.KG

Device : Gammapilot M

Type : FMG60

Previous software version : V01.02.00 4...20mA HART

How can the previous software version number be identified:

- (a) by standard HART Comando #0
- (b) by ToF Tool
- (c) by local operation FHX40 (Group selection-> system parameters
-> protocol+sw-no.
by FieldCare

2 Firmware/Software

New firm-/software version : V01.02.02 4...20mA HART

Description of the modification in comparison with the predecessor version:

- (1) Max. edit value at density from 16 000 c/s to 160 000 c/s
- (2) HV controller optimized

3 Compatibility

Is the newly installed firmware/software compatible with the operating tools?

- ☒ Yes
 - ToF Tool Field Tool Package version 4.04 or higher
 - HART Communicator DXR375
 - FHX40 (VU331)
 - FieldCare from V 2.03.00 + DTM Library V 2.15.00

☐ No, description:

Manufacturer Information

for users regarding software updates
(following the NAMUR recommendation 53)

Is a firm-/software update generally recommended?

☐ Yes, reasons:

☒ No, reasons:

During automatic calibration the max. value from 160 000 c/s is possible. Limitation only at manual value edits.

HV controller optimized for locations with low radiation level. Only if error message W536 or A535 during commissioning appear is software update necessary.

4 Instruction manual

Is a new instruction manual necessary due to the modification of the firm-/software?

☐ Yes

☒ No

Which manual corresponds to the new firm-/software:

Device	Communication options	Manual	Marking
FMG60	4...20mA HART	Instruction manual	BA236F/00/en/06.07 71041168
FMG60	HART, PA, FF	Description of Instrument Functions	BA287F/00/en/06.07 on ToF-Tool-CD
FMG60		Functional safety manual	SD230F/00/en/10.07 No. 71041846

The new instruction manuals can be referred to in Internet:

<http://www.endress.com>

- area „DOWNLOAD“

- declaration of the device and kind of manual

5 Price

Change in price of device in comparison with the predecessor version?

☐ Yes, new list price and update costs (without installation) are enclosed

☒ No