



Certificate of Compliance

Certificate: 70157089

Master Contract: 205557

Project: 70173936

Date Issued: 2018-03-06

Issued to: Endress+Hauser Conducta GmbH & Co. KG
Dieselstrasse 24
Gerlingen, 70839
GERMANY
Attention: Robert Binder

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: *Dimcho Genov*
Dimcho Genov

PRODUCTS

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT-Certified to US Standards

Two-wire transmitter for process engineering, Models Liquiline Compact CM82, Liquiline Compact CM72

Rated 30VDC, 23mA; ambient rated -20°C to +85°C; Enclosure type 6

Notes:

1. To be supplied by a Class 2 or Limited Energy Source in accordance with CSA 61010-1-12.
2. Overvoltage Category I.
3. Environmental Conditions: 2000 m max

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

ANSI/UL Standard 61010-1

- Electrical Equipment for Measurement, Control, and Laboratory Use; Part 1: General Requirements (Third Edition)



Supplement to Certificate of Compliance

Certificate: 70157089

Master Contract: 205557 (205557)

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70173936	2018-03-06	Update report 70157089 to cover an addition of an alternate housing material.
70157089	2017-11-28	New Certification of models CM82 and CM72