

# Herstellererklärung Manufacturer Declaration

**Company**                      **Endress+Hauser Flowtec AG, Kägenstrasse 7, CH-4153 Reinach**  
declares as manufacturer, that the following product

**Product**                      **Promass U 500**

| Description           | Order code    | Extended order code | Mat.Nr.  |
|-----------------------|---------------|---------------------|----------|
| Promass U DN25 / 1"   | DK8014-10D8/0 | DK8014-25SBOADFA2   | 70212574 |
| Promass U DN15 / 1/2" | DK8014-10C2/0 | DK8014-15SBOACFA2   | 70212573 |
| Promass U DN06 / 1/4" | DK8014-10A1/0 | DK8014-06SBOABFA2   | 70212572 |
| Promass U DN04 / 1/8" | DK8014-1098/0 | DK8014-04SBOAAFA2   | 70212568 |

This manufacturer declaration is exclusively valid for the listed products in delivery status.

The document describes the Certificate of Compliance (CoC) for biopharmaceutical single-use requirements.

CoC Document ID 6407806


The above generated document ID will be generated electronically and is valid without signature.

This manufacturer declaration is exclusively valid for the customer listed in the cover letter of the respective Endress+Hauser sales center and for the listed products in delivery status.

The validity of this manufacturer declaration expires 2 years after the date of issue.

Reinach, 17.5.2024  
Endress+Hauser Flowtec AG

  
Dr. Mirko Lehmann  
Managing Director

  
Dr. Christian Jarms  
Head of Division Quality Management

## Certificate of Compliance (CoC) for biopharmaceutical single-use requirements.

### Compliance to ASME BPE 2022

Hereby we confirm that the product supplied is in compliance to all relevant parts of ASME BPE latest revision (2022) and with the requirements of the order. Furthermore, we declare that during the manufacturing of the product supplied, the valid Endress+Hauser procedures have been followed. Non-specific tests and inspections have been performed and the relevant releases have been given.

### Materials of construction of product wetted parts

The product mentioned above consists of process wetted components listed in the table below.

### TSE/BSE compliance

All wetted parts do not derive from animal sources and comply with the regulatory requirements of guidance EMA/410/01 Rev. 3.

Furthermore, no grinding and polishing agents of animal origin have been used during the entire production process.

### Compatibility with sterilization operations

The materials selected withstand the following sterilization modes without impairing functional performance. Gamma/e-beam irradiation up to 50 kGy (5.0 Mrad), steam/autoclave 121 °C (250 °F) 60 min (1 cycle).

### Assembly and Packaging

Assembly and packaging of the device were carried out in an FDA registered (Samaplast AG registration number # 3008793310) and ISO 13485 certified facility within an ISO 14644-1 Class 7 cleanroom manufacturing suite.

### Polishing and surface finish table of process wetted parts

| Component      | Surface finish Ra <sub>max</sub> | Polishing <sup>1</sup> | ASME BPE designation           |
|----------------|----------------------------------|------------------------|--------------------------------|
| Measuring tube | ≤ 0.76 µm (30 µin)               | mechanically polished  | Acc. to table SF-2.4.1-1 (SF3) |
| Flow splitter  | ≤ 0.76 µm (30 µin)               | n.a.                   | Acc. to table SF-3.4-1 (SFP3)  |

<sup>1</sup> No polishing compound has been used.

### Material / compound

| Component         | E+H Part number    | Wetted materials | Material / compound            |
|-------------------|--------------------|------------------|--------------------------------|
| Seal O-ring       | 71580323           | yes              | Silicone VMQ                   |
| Measuring tube    | 71535529           | yes              | Stainless steel 1.4435 (316L)  |
| Flow splitter     | 71535528           | yes              | Polycarbonate Makrolon® Rx1805 |
| Double pouch bags | 71597954, 71597955 | yes              | PET-OPA-PE-PEEL                |
| Blister           | 71580443           | yes              | PET                            |

## Regulatory / specification

| Requirement                           | Standard                             | Components  |                |               |                   |         |                  |
|---------------------------------------|--------------------------------------|-------------|----------------|---------------|-------------------|---------|------------------|
|                                       |                                      | Seal O-ring | Measuring tube | Flow splitter | Double pouch bags | Blister | Assembled device |
| Bioburden                             | ISO 11737-1 or USP<788> or EP-2.9.19 | X           | X              | X             | X                 | X       | X                |
| Cleanliness                           | ISO 19227:2018                       | X           | X              | X             | X                 | n.a.    | X                |
| Cytotoxicity                          | ISO 10993-5 or USP<42>, 2019 Ch. 87  | X           | n.a.           | X             | X                 | X       | X                |
| Medical packaging                     | EN ISO 11607-1                       | n.a.        | n.a.           | n.a.          | X                 | X       | n.a.             |
| Medical packaging                     | ISTA-3A                              | n.a.        | n.a.           | n.a.          | n.a.              | n.a.    | X                |
| Endotoxin <sup>2</sup>                | USP<85> or EP 2.6.14                 | n.a.        | n.a.           | n.a.          | n.a.              | n.a.    | X                |
| Sterilization validation <sup>2</sup> | ISO 11737-2 or USP<71>               | n.a.        | n.a.           | n.a.          | n.a.              | n.a.    | X                |
| FDA compliance                        | FDA 21 CFR 177.2600 or EU 1935/2004  | X           | n.a.           | n.a.          | X                 | X       | n.a.             |

<sup>2</sup> Only applicable after gamma irradiation.

## Biocompatibility compliance table

| Requirement              | Standard                        | Components  |                |               |                   |         |                  |
|--------------------------|---------------------------------|-------------|----------------|---------------|-------------------|---------|------------------|
|                          |                                 | Seal O-ring | Measuring tube | Flow splitter | Double pouch bags | Blister | Assembled device |
| Sensitization            | ISO 10993-10                    | n.a.        | X              | X             | X                 | X       | n.a.             |
| Systemic toxicity        | ISO 10993-11                    | X           | X              | X             | n.a.              | n.a.    | n.a.             |
| Sample preparation       | ISO 10993-12                    | X           | n.a.           | n.a.          | X                 | X       | n.a.             |
| Reactivity test in vivo  | ISO 10993-1 or USP<88> Class VI | X           | n.a.           | X             | X                 | X       | X                |
| Reactivity test in vitro | ISO 10993-5 or USP<87>          | X           | n.a.           | X             | X                 | X       | X                |

## Hemolyses & extractable data compliance table

| Requirement                  | Standard | Components  |                |               |                   |         |                  |
|------------------------------|----------|-------------|----------------|---------------|-------------------|---------|------------------|
|                              |          | Seal O-ring | Measuring tube | Flow splitter | Double pouch bags | Blister | Assembled device |
| Physical chemical resistance | USP<661> | n.a.        | n.a.           | X             | X                 | n.a.    | n.a.             |