

UNITED KINGDOM CONFORMITY ASSESSMENT

UK TYPE EXAMINATION CERTIFICATE 1

2 Equipment Intended for use in Potentially Explosive Atmospheres

UKSI 2016:1107 (as amended) - Schedule 3A, Part 1

Certificate Number: CSAE 21UKEX1072X 3 Issue:

J22 TDLAS Gas Analyzer, J22 TDLAS Gas Analyzer SCS on Panel, J22 4 Product:

TDLAS Gas Analyzer Encl. SCS, J22 TDLAS Gas Analyzer Encl. SCS Heated

5 Manufacturer: Endress+Hauser Optical Analysis Inc.

Address: 11027 Arrow Route, Rancho Cucamonga, California, 91730, USA 6

7 This product and any acceptable variation thereto is specified in the schedule to this certificate and the documents therein referred to.

- CSA Group Testing UK Limited, Approved Body number 0518, in accordance with Regulation 42 of the 8 Equipment and Protective Systems Intended for Use in Potentially Explosive Atmospheres Regulations 2016, UKSI 2016:1107 (as amended), certifies that this product has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of products intended for use in potentially explosive atmospheres given in Schedule 1 of the Regulations. The examination and test results are recorded in the confidential reports listed in Section 14.2.
- 9 Compliance with the Essential Health and Safety Requirements has been assured by compliance with:

EN IEC 60079-0:2018 EN 60079-1:2014+AC:2018-09 EN 60079-11:2012

FN 60079-28:2015 EN ISO 80079-36:2016+AC:2019 IFC TS 60079-40:2015 Edition 1.0

Except in respect of those requirements listed at Section 16 of the schedule to this certificate. The above standards may not appear on the UKAS Scope of Accreditation, but have been added through flexible scope of accreditation, which is available on request.

- 10 If the sign 'X' is placed after the certificate number, it indicates that the product is subject to Specific Conditions of Use identified in the schedule to this certificate.
- 11 This UK TYPE EXAMINATION CERTIFICATE relates only to the design and construction of the specified product. Further requirements of the Regulations apply to the manufacturing process and supply of this product. These are not covered by this certificate.
- 12 The marking of this product shall be in accordance with Regulation 41 and include the following:

J22 TDLAS Gas Analyzer



II 2G Ex db ia [ia Ga] ib op is IIC T4 Gb $-20^{\circ}\text{C} \le \text{Ta} \le +60^{\circ}\text{C}$

J22 TDLAS Gas Analyzer Encl. SCS Heated



II 2G Ex db ia ib op is h IIC T3 Gb $-20^{\circ}C \le Ta \le +60^{\circ}C$

J22 TDLAS Gas Analyzer SCS on Panel or J22 TDLAS Gas Analyzer Encl. SCS



II 2G

Ex db ia ib op is h IIC T4 Gb $-20^{\circ}C \le Ta \le +60^{\circ}C$



Title: Director of Operations





UK TYPE EXAMINATION CERTIFICATE

CSAE 21UKEX1072X Issue 4

13 DESCRIPTION OF PRODUCT

The core model of the J22 TDLAS Gas Analyzer consists of a flameproof electronics compartment, intrinsically safe optical head and a measurement cell.

The J22 TDLAS Gas Analyzer with Sample Conditioning System, henceforth referred to as the 'SCS', on a panel combines the J22 TDLAS Gas Analyzer with the non-electrical SCS to precondition the analyte before examination within the cell.

The J22 TDLAS Gas Analyzer with SCS in an enclosure can be configured with or without a pre-certified terminal box, heater and thermostat.

The J22 TDLAS Gas Analyzer is a laser-based gas analyzer that measures the concentration of a compound or "analyte" such as H2O. The technology employed is Tunable Diode Laser Absorption Spectroscopy (TDLAS). The measurement output is a volumetric concentration, or ratio of a specific analyte in a gas mixture such as natural gas or air. The volumetric ratio can be converted to other units of measure using internal conversion factors and calculations.

The analyzer is comprised of a sample cell, intrinsically safe optical head and an electronics assembly platform within a pre-certified flameproof enclosure. The cell is a sealed tube through which the gas mixture flows. The cell has a gas inlet and a gas outlet. On one end of the tube is a window through which a beam of infrared laser light travels, which in turn reflects from a mirror. In this arrangement, the gas mixture does not contact the laser or any other optoelectronics. Pressure, and in some cases temperature sensors, are employed in the cell assembly to compensate for the effects of pressure and temperature changes in the gas.

The optical head is mounted on top of the cell and contains the laser, optical detector and a thermoelectric cooler to control the laser temperature. The optical head also contains the optical head electronics which are directly connected to the optoelectronics in the optical head. The optical head electronics board also communicates with the electronics assembly.

The electronics assembly is mounted on top of the optical head, within a flameproof enclosure. The electronics assembly, which can be powered by 100-240 VAC \pm 10% or 19.2-28.8 VDC source, contains the sensor electronics which connects to the optical head via the RS232 protocol through a 10-pin ribbon cable assembly. The sensor electronics and the optical head electronics operate on a 30V dc supply using the same 10-pin ribbon cable. The sensor electronics generate the laser drive signal that is sent through the optical head electronics and to the laser in the optical head. Signals from the detectors are amplified by the optical head electronics and sent to the sensor electronics where they are digitized. The sensor electronics process the digital data and sends the gas concentration measurements the electronics display and I/O modules.

The electronics assembly displays the concentration measurement on an LCD display and also has a through-the-glass 3-button keypad interface for user input. The flameproof enclosure of the electronics assembly also houses the electrical terminals for field wiring connections. The J22 comes with various analogue and digital outputs which may be employed in automation or communication systems to deliver its measurements and applicable diagnostic messages and alarms to remote devices. Additionally, the electronics assembly has a Service Port which allows interaction with the J22 TDLAS Gas Analyzer on a standard web browser using a laptop or tablet. This connection is for use by the manufacturer or trained personnel for test, repair or overhaul interaction of the equipment under non-hazardous, explosive atmosphere free conditions.





UK TYPE EXAMINATION CERTIFICATE

CSAE 21UKEX1072X Issue 4

'Within the optical head enclosure there is an intrinsically safe 'flow switch' connection for connection of external equipment, and an intrinsically safe RS485 interface for future use by the manufacture for connection to additional equipment that it is intended to be incorporated as part of the SCS.'

The J22 TDLAS Gas Analyzer is used as an "extractive" measuring device, where the gas sample is extracted from a vessel or pipeline and transported to the analyzer which may be mounted up to 100m from the sample tap point. The J22 TDLAS Gas Analyzer may be configured without sample conditioning. Optional hardware is also available to condition the sample before it enters the gas analyzer. An optional IP66/Type 4X rated enclosure may be included which surrounds the "cell" and the sample conditioning system. The sample conditioning system (SCS) is required to filter and remove solid and liquid particulate and to control the flow and pressure of the gas. In some cases, an optional heater (powered independently from the analyzer) is used to control the temperature inside the enclosed SCS. An optional pressure relief valve and purge system are available. The effectiveness of this purge system has not been assessed by CSA.

The J22 TDLAS Gas Analyzer operates at near-atmospheric pressure and at an ambient temperature between negative -20 to +60 degrees centigrade. After the passing through the J22 TDLAS Gas analyzer, the sample is routed and vented to a safe location in the atmosphere or vented to a flare or other apparatus.

The equipment has been separately tested against the requirements of IEC 60529 and it meets IP66.

The maximum working pressure (MWP) of the equipment is listed as 800-1200 mbara, or 800-1700 mbara (model dependant). This is the pressure range at which the manufacturer has determined the equipment can operate. However, the equipment was assessed to withstand a pressure of 100 lbf/in² (psi) against IEC TS 60079-40:2015 Edition 1.0.

J22 TDLAS Gas Analyzer

Rated: 100 - 240 Vac, $50/60 \text{ Hz} \pm 10\%$, Um = 250 V or 19.2 - 28.8 Vdc, max., $\text{Um} \ 250 \text{V}$, 10 W.

I/01: Terminal 26 and 27, Un = 30Vdc, Um = 250Vac

I/02: Terminal 24 and 25, Un = 30Vdc, Um = 250Vac or Un = 30Vdc, In = 100mAdc/500mAac, Um = 250Vac

I/03: Terminal 22 and 23, Un = 30Vdc, Um = 250Vac or Un = 30Vdc, In = 100mAdc/500mAac, Um = 250Vac

Flow Switch: J6 (Optical Head Enclosure), Uo = 5.88V, Io = 4.53mA, Po = 6.66mW, Co = $43\mu F$, Lo = 1.74H (Uo may be + or – 5.88V with respect to Pin 2 of J6)'

J22 TDLAS Gas Analyzer SCS on Panel & J22 TDLAS Gas Analyzer Encl. SCS

Rated: 100 - 240 Vac, $50/60 \text{ Hz} \pm 10\%$, 100 Um = 250 V or 19.2 - 28.8 Vdc, 100 Um = 250 V, 100 Um = 250 V.

I/01: Terminal 26 and 27, Un = 30Vdc, Um = 250Vac

I/02: Terminal 24 and 25, Un = 30Vdc, Um = 250Vac or Un = 30Vdc, In = 100mAdc/500mAac, Um = 250Vac

I/03: Terminal 22 and 23, Un = 30Vdc, Um = 250Vac or Un = 30Vdc, In = 100mAdc/500mAac, Um = 250Vac

J22 TDLAS Gas Analyzer Encl. SCS Heated

Rated: 100 - 240 Vac, $50/60 \text{ Hz} \pm 10\%$, 100 Um = 250 V or 19.2 - 28.8 Vdc, 100 Um = 250 V, 100 Um = 250 V.

Heater: 100 - 240 Vac, 50/60 Hz ± 10%, 80 W.

I/01: Terminal 26 and 27, Un = 30Vdc, Um = 250Vac

 $I/02: Terminal\ 24\ and\ 25,\ Un=30Vdc,\ Um=250Vac\ or\ Un=30Vdc,\ In=100mAdc/500mAac,\ Um=250Vac$

I/03: Terminal 22 and 23, Un = 30Vdc, Um = 250Vac or Un = 30Vdc, In = 100mAdc/500mAac, Um = 250Vac





UK TYPE EXAMINATION CERTIFICATE

CSAE 21UKEX1072X Issue 4

Model Code Structure

J22 – ABCDEFGHIJKLMNOPQRSTUVW. The manufacturer's internal reference code is shown adjacent to each model code placeholder in brackets ().

A (10) - Approval

BA - ATEX/IECEx/UKEx Zone 1

- B (20) Analyte
- C (30) Measurement Range
- D (40) Measurement Range 2
- E (50) Stream Composition
- F (60) Venting to
- G (70) Process Wetted Materials
 - J 316 Stainless Steel; FFKM Seals (Analyzer Only)
 - K 316 Stainless Steel; FFKM Seals (Analyzer with SCS)
 - V 316 Stainless Steel; FKM Seals
- H (80) Supply Parameters
 - A $100-240 \text{ VAC } (50/60 \text{ Hz}) \pm 10\%$
 - D-24 VDC \pm 20%
- I (90) Output; Input 1
- J (100) Output; Input 2
- K (110) Output; Input 3
- L (120) Electronics Housing
 - 1 Coated Copper-Free Aluminum; no WLAN
 - 2 316 Stainless Steel; no WLAN
 - 3 Coated Copper-Free Aluminum + WLAN
 - 4 316 Stainless Steel + WLAN
- M (130) Controller Mounting
- N (140) Sample Conditioning System (SCS)
 - A On Panel, Aluminum
 - B Enclosed, 304 Stainless Steel
 - C Enclosed, 316 Stainless Steel
 - N None
- O (150) Filtration
- P (160) Sample System Gas Connections
 - A Imperial
 - B Metric
- Q (170) Pressure Regulation
- R (180) Flow Meter
 - G Armored, factory default
 - P Armored, Krohne with flow switch
 - F Glass Tube, factory default
 - K Glass Tube, Krohne
 - N None
- S (190) Heating Options
 - 1 Heated + Heat-Trace Boot, 100 240 VAC ± 10%
 - 8 None
- T (200) Purge





UK TYPE EXAMINATION CERTIFICATE

CSAE 21UKEX1072X Issue 4

U (500) - Operating Language Display

V (580) - Test/Certificate/Declaration

W (895) - Marking

Headings without sub-options are not considered critical to the design of the equipment. Where sub options are shown, these are the only options endorsed by CSA.

Variation 1 - This variation introduced the following changes:

- i. Change of manufacturer name from SpectraSensors Inc. to Endress+Hauser Optical Analysis Inc.
- ii. Specification of intrinsic safety 'Ex ia' parameters at a previously unused flow switch connector J6 of the J22 TDLAS Gas Analyzer to cover user connection of intrinsically safe equipment at this connector. The addition of 'ia [ia Ga]' to the certification coding to cover this connection and its associated circuitry.
- iii. Introduction of the option for an intrinsically safe 'Ex ia' flowmeter with flow switch to be fitted as part of the Sample Conditioning System, the flow switch of the flowmeter being electrically connected to the previously unused intrinsically safe 'Ex ia' connection at connector J6 in the J22 TDLAS Gas Analyzer. The addition of 'ia' to the certification coding to cover the fitting of this flowmeter with flow switch.
- iv. The use of alternative internal p.c.b. assemblies, including an alternative terminal board assembly, to provide Modbus Ethernet/IP at I/O1 as an optional alternative to the Modbus RS485 at I/O1.
- v. Clarification that internal board configurations and setting can optionally provide either Relay Output or 4-20mA Input/Output (Passive/Active) at I/O2 I/O3.
- vi. Modifications to the design of the TLM Module.
- vii. Minor modifications to an internal pressure sensor.
- viii. Introduction of alternative non-return valves Swagelock SS-CHS4-1/3 and DK-LOK CORP. V33A-D-4T-1/3 to the Sample Conditioning System.

Variation 2 - This variation introduced the following changes:

- i. The equipment was assessed against the requirements of IEC TS 60079-40:2015. As a result of this assessment, the marking "Dual Seal Without Annunciation" was added.
- ii. Minor updates were made to the nomenclature in the Model Code Structure, listed under the Product Description.
- iii. Minor updates were made to the manufacturer's controlled documents.

Variation 3 - This variation introduced the following changes:

- i. Recognition of a RS485 interface for future use by the manufacture for connection to additional equipment that is intended to be incorporated as part of the SCS. The Product Description is updated accordingly.
- ii. Minor updates to the assembly drawing.
- iii. Removal of the superfluous words 'Dual Seal Without Annunciation' from the product name in the marking section of the certificate.
- v. Minor updates were made to the manufacturer's controlled documents.

Variation 4 - This variation introduced the following changes:

- i. Addition of alternative lens material.
- ii. Addition of alternative "Kalrez" O-rings in the process wetted area.
- iii. Addition of Wi-Fi Display.
- iv. Addition of alternative pressure relief valves and check valves.





UK TYPE EXAMINATION CERTIFICATE

CSAE 21UKEX1072X Issue 4

- v. Update of component certificates for E+H pre-certified parts used in the equipment.
- vi. Addition of stainless steel H328 adapter.
- vii. As a result of the above modifications, the Product Description and controlled drawings were updated as appropriate.

14 DESCRIPTIVE DOCUMENTS

14.1 Drawings

Refer to Certificate Annexe.

14.2 Associated Reports and Certificate History

| Issue | Date | Report number | Comment |
|-------|------------------|---------------|---------------------------------------|
| 0 | 09 July 2021 | R80074463A | The release of the prime certificate. |
| 1 | 28 February 2022 | R80099539A | The introduction of Variation 1. |
| | - | R80099539B | |
| 2 | 27 October 2022 | R80132099A | The introduction of Variation 2. |
| 3 | 04 January 2023 | R80125323A | The introduction of Variation 3. |
| 4 | 12 December 2024 | R80226186A | The introduction of Variation 4. |
| | | R80226186B | |

- 15 SPECIFIC CONDITIONS OF USE (denoted by X after the certificate number)
- 15.1 The flameproof joints of this equipment are other than the minimums specified in IEC/EN 60079-1 and shall not be repaired by the user.
- 15.2 Adhesive labels, the powder coating of models of the equipment with an aluminium enclosure, and coated parts of the flowmeter with flow switch are non-conducting materials and may generate an ignition-capable level of electrostatic discharge under certain extreme conditions. The user should ensure that the Equipment is not installed in a location where it may be subjected to external conditions (such as high-pressure steam) which might cause a build-up of electrostatic charges on these non-conducting surfaces. Additionally, cleaning of the equipment should be done only with a damp cloth.
- 15.3 The optional stainless-steel label tag is not bonded to earth. The maximum average capacitance of the tag determined by measurement is max. 30 pF. This shall be considered by the user to determine suitability of the equipment in a specific application.
- For models of the J22 TDLAS Gas Analyzer with SCS mounted within an enclosure, the inner sheath of the supply cable for the heater circuit shall be sheathed with thermoplastic, thermosetting, or elastomeric material. It shall be circular and compact. Any bedding or sheath shall be extruded. Fillers, if any, shall be non-hygroscopic. The minimum length of the cable shall exceed 3 meters.
- 15.5 The temperature of the process medium shall be within the ambient temperature rating of the equipment.
- 15.6 The equipment is not capable of passing a 500V r.m.s. dielectric strength test in accordance with Clause 6.3.13 of EN IEC 60079-11:2012 between the Intrinsically Safe Flow Switch connection circuits and the equipment enclosure. This shall be taken into account in any equipment installation.
- 15.7 This equipment is intended to operate at a constant pressure and has not been assessed for the effects of persistent fluctuations of pressure within the operating pressure range. Therefore, the user shall





UK TYPE EXAMINATION CERTIFICATE

CSAE 21UKEX1072X Issue 4

ensure that the pressure fluctuation within the Sample Cell Tube of the equipment does not routinely exceed 5 lbf/in² (psi).

16 ESSENTIAL HEALTH AND SAFETY REQUIREMENTS (REGULATIONS SCHEDULE 1)

In addition to the Essential Health and Safety Requirements covered by the standards listed in Section 9, all other requirements are demonstrated in the relevant reports.

17 PRODUCTION CONTROL

- 17.1 Holders of this certificate are required to comply with production control requirements defined in Schedule 3A, as applicable, and CSA Group Testing UK Regulations for Certificate Holders
- 17.2 The equipment covered by this certificate incorporates previously certified devices; it is therefore the responsibility of the manufacturer to continually monitor the status of the certification associated with these devices, and the manufacturer shall inform CSA UK of any modifications of the devices that may impinge upon the explosion safety design of the equipment.

| Description | Certificate Number |
|---|------------------------------------|
| E+H G305 and G307 Enclosures | IECEx CSA 20.0016U |
| Intertec SL-***THERM Block Heater | PTB 02 ATEX 1116 X |
| Intertec TA Thermostat | PTB 03 ATEX 1136 X |
| Adalet XIHS Terminal Box | DEMKO 12 ATEX 1115099U |
| CMP Type 737 Adapter/Reducer | CML 18ATEX1320X |
| CMP A2F Cable Gland | CML 18ATEX1321X IECEx CML 18.0179X |
| Hazardous Locations Solutions Type N conduit elbow | IECEx SIR 07.0044U |
| Proline 300/500 Electronics | IECEx CSA 16.0006U |
| Display Type DP-PA1c & DP-PH1c | IECEx KEM 08.0048U |
| KROHNE Messtechnik GmbH | KIWA 18ATEX0008 X |
| Variable Area Flowmeter type DK32 / R1 / / L / / – Ex | |





Certificate Annexe

Certificate Number: CSAE 21UKEX1072X

Product: J22 TDLAS Gas Analyzer, J22 TDLAS Gas Analyzer SCS on Panel,

J22 TDLAS Gas Analyzer Encl. SCS, J22 TDLAS Gas Analyzer Encl.

SCS Heated

Manufacturer: Endress+Hauser Optical Analysis Inc.

Issue 0

| Drawing | Sheets | Rev. | Date (Stamp) | Title |
|-----------------|---------|------|--------------|--|
| 4900002311 | 1 to 4 | С | 16 Feb 21 | J22 Analyzer Product Description Document |
| MPI-100025 | 1 of 1 | Α | 16 Feb 21 | Manufacturing Process Instruction Potting Process RT |
| | | | | 622, Exd and CLASS-1 DIVISION-1 XP |
| EX090000012 | 1 to 2 | E | 16 Feb 21 | Housing, Conn Stem, H328 |
| EX090000025 | 1 to 2 | С | 16 Feb 21 | Enclosure, Optical Head |
| EX090000026 | 1 of 1 | В | 16 Feb 21 | Cover, Opt Head Enclosure |
| EX090000028 | 1 of 1 | Α | 16 Feb 21 | Hot Plate, TLM |
| EX090000029 | 1 of 1 | Α | 16 Feb 21 | Cold Plate, TLM |
| EX090000030 | 1 of 1 | Α | 16 Feb 21 | Cover, Opt Block |
| EX090000032 | 1 of 1 | Α | 16 Feb 21 | Opt block, WW, DD, Cert |
| EX1300000038 | 1 of 1 | В | 16 Feb 21 | Gasket, H328 Collar / Opt Enclsr |
| EX1300000039 | 1 of 1 | В | 16 Feb 21 | Gasket, Enclosure Access |
| EX170000004 | 1 to 6 | D | 07 Jun 21 | Label, J22 Rating |
| EX2100000002 | 1 to 2 | В | 16 Feb 21 | Bushing + 28 Contacts |
| EX210000010 | 1 to 2 | Α | 16 Feb 21 | Conn, 28-Pin, 90V Isolation |
| EX290000010 | 1 to 4 | D | 07 Jun 21 | PCBA, Optical Head Assembly |
| EX290000010-00 | 1 to 3 | E | 07 Jun 21 | BOM, Optical Head Electronics |
| EX290000011 | 1 to 12 | В | 07 Jun 21 | PCB, Optical Head Fabrication Notes |
| EX290000012 | 1 to 9 | С | 07 Jun 21 | Schem, Optical Head Electronics |
| EX290000020 | 1 to 4 | E | 07 Jun 21 | PCBA, ISEM Analog Board Assembly |
| EX2900000020-00 | 1 to 3 | E | 07 Jun 21 | BOM, ISEM Analog |
| EX2900000021 | 1 to 14 | С | 07 Jun 21 | PCB, ISEM Analog Board Fabrication Notes |
| EX290000022 | 1 to 9 | С | 07 Jun 21 | Schem, ISEM Analog |
| EX290000030 | 1 to 4 | D | 07 Jun 21 | PCBA, ISEM Digital MCU Assembly |
| EX290000030-00 | 1 to 2 | D | 07 Jun 21 | BOM, ISEM Digital MCU Board |
| EX290000031 | 1 to 16 | В | 07 Jun 21 | PCB, ISEM Digital MCU Fabrication Notes |
| EX290000032 | 1 to 17 | В | 07 Jun 21 | Schem, ISEM Digital MCU |
| EX290000040 | 1 to 3 | Α | 16 Feb 21 | PCBA, Transition-Laser/Detector Assembly Side-A |
| EX2900000040-00 | 1 to 2 | Α | 16 Feb 21 | BOM, PCBA Assy, Transition – Laser/Detector |
| EX2900000041 | 1 to 10 | Α | 16 Feb 21 | PCB, Transition-Laser/Detector Fabrication Notes |
| EX2900000042 | 1 to 3 | Α | 16 Feb 21 | Schem, Transition – Laser/Detector |
| EX290000071 | 1 to 5 | В | 16 Feb 21 | PCB, TLM Flex Fabrication Notes |
| EX290000072 | 1 to 3 | В | 16 Feb 21 | Schem, TLM Flex |
| EX2900000081 | 1 to 7 | С | 16 Feb 21 | PCB, ISEM 28 Flex, Analog Fabrication Notes |
| EX2900000082 | 1 to 3 | Α | 16 Feb 21 | Schem, ISEM 28 Flex, Analog |
| EX2900000091 | 1 to 7 | С | 16 Feb 21 | PCB, ISEM 28 Flex, Digital Fabrication Notes |
| EX290000092 | 1 to 3 | Α | 16 Feb 21 | Schem, ISEM 28 Flex, Digital |





Certificate Annexe

Certificate Number: CSAE 21UKEX1072X

Product: J22 TDLAS Gas Analyzer, J22 TDLAS Gas Analyzer SCS on Panel,

J22 TDLAS Gas Analyzer Encl. SCS, J22 TDLAS Gas Analyzer Encl.

SCS Heated

Manufacturer: Endress+Hauser Optical Analysis Inc.

| Drawing | Sheets | Rev. | Date (Stamp) | Title |
|--------------|---------|------|--------------|---|
| EX2900000141 | 1 to 3 | Α | 16 Feb 21 | PCB Heat Spreader TLM (2x5) |
| EX2900000161 | 1 to 4 | Α | 16 Feb 21 | PCB Flex, PDMx1 |
| EX2900000162 | 1 to 3 | Α | 16 Feb 21 | Schematic, PCB FLEX, PDMx1 |
| EX2900000171 | 1 to 4 | Α | 16 Feb 21 | PCB, ISEM Shield |
| EX5500000004 | 1 to 2 | В | 16 Feb 21 | Pressure Sensor, Digital |
| EX600000009 | 1 of 1 | Α | 16 Feb 21 | Cable, P3 to ISE MCU Digital Brd |
| EX600000012 | 1 of 1 | Α | 16 Feb 21 | Cable Assy, Pressure, Digital |
| EX800000062 | 1 to 3 | С | 16 Feb 21 | Isem Interface Module Assy |
| EX800000092 | 1 to 3 | E | 16 Feb 21 | Analyzer Assy, 0.8m, Unit Sealing |
| EX800000093 | 1 to 4 | С | 16 Feb 21 | Analyzer System Assy, Unit Sealing |
| EX800000104 | 1 of 1 | Α | 16 Feb 21 | Cover Assy, Opt Block |
| EX800000107 | 1 to 2 | С | 18 Jun 21 | Det Module, Parabolic, TO18, Sig |
| EX8000000111 | 1 to 2 | Α | 16 Feb 21 | Optical Head Assy, Cert |
| EX800000112 | 1 to 3 | D | 16 Feb 21 | Analyzer Assy, Cert |
| EX800000113 | 1 to 5 | Α | 16 Feb 21 | TLM ASSY, Cert |
| EX8000000114 | 1 to 11 | Α | 16 Feb 21 | SCS Assy, Enclosure, J22, Certification |
| EX870000015 | 1 of 1 | Α | 16 Feb 21 | T-Plate Assy, 0.8m |

Issue 1

| Drawing | Sheets | Rev. | Date (Stamp) | Title |
|-------------|---------|------|--------------|---|
| EX090000028 | 1 of 1 | В | 02 Feb 22 | Hot Plate, TLM |
| EX090000029 | 1 of 1 | В | 02 Feb 22 | Cold Plate, TLM |
| EX170000004 | 1 to 6 | G | 16 Feb 22 | Label, J22 Rating |
| EX290000071 | 1 to 5 | С | 02 Feb 22 | PCB, TLM Flex Fabrication Notes |
| EX550000004 | 1 and 2 | D | 16 Feb 22 | Pressure Sensor, Digital |
| EX800000055 | 1 of 1 | Α | 02 Feb 22 | Hot Plate – TEC Asst, TLM |
| EX800000055 | 1 of 1 | С | 09 Feb 22 | Hot Plate – TEC Asst, TLM |
| EX800000092 | 1 to 3 | F | 09 Feb 22 | Analyzer Assy, 0.8M, Unit Sealing |
| EX800000112 | 1 to 4 | E | 09 Feb 22 | Analyzer Assy, Cert |
| EX800000113 | 1 to 5 | С | 09 Feb 22 | TLM ASSY, Cert |
| EX800000114 | 1 to 11 | С | 11 Feb 22 | SCS Assy, Enclosure, J22, Certification |

Issue 2

| Drawing | Sheets | Rev. | Date (Stamp) | Title |
|-----------------|--------|------|--------------|--------------------------|
| EX1700000004-BA | 1 to 5 | Α | 25 Oct 22 | Label, J22 Rating, ATEX |
| EX870000023 | 1 to 1 | Α | 09 Sep 22 | Window (optical head) |
| EX5500000004 | 1 to 2 | E | 09 Sep 22 | Pressure Sensor, Digital |
| EX870000015 | 1 to 1 | В | 25 Oct 22 | T-Plate Assy, 0.8m |





Certificate Annexe

Certificate Number: CSAE 21UKEX1072X

Product: J22 TDLAS Gas Analyzer, J22 TDLAS Gas Analyzer SCS on Panel,

J22 TDLAS Gas Analyzer Encl. SCS, J22 TDLAS Gas Analyzer Encl.

SCS Heated

Manufacturer: Endress+Hauser Optical Analysis Inc.

Issue 3

| Drawing | Sheets | Rev. | Date (Stamp) | Title |
|-------------|---------|------|--------------|---|
| EX800000114 | 1 to 11 | D | 03 Jan 23 | SCS ASSY, Enclosure, J22, Certification |

Issue 4

| Drawing | Sheets | Rev. | Date (Stamp) | Title |
|----------------|---------|------|--------------|---|
| EX170000004-BA | 1 to 5 | В | 18 Nov 24 | LABEL, J22 RATING, ATEX |
| EX090000017 | 1 to 1 | Α | 25 Feb 24 | Adapter, H328, SS, Cert |
| EX3700000021 | 1 to 1 | Α | 18 Nov 24 | WINDOW,WDG 1.87°, Ø25.4x2.0mm,1700-2030nm |
| EX5500000004 | 1 to 2 | F | 14 Aug 24 | PRESSURE SENSOR, DIGITAL |
| EX5500000008 | 1 to 2 | Α | 14 Aug 24 | PRESSURE SENSOR, DIGITAL, KALREZ |
| EX8000000111 | 1 to 2 | В | 18 Nov 24 | OPTICAL HEAD ASSY, CERT |
| EX800000112 | 1 to 4 | G | 18 Nov 24 | ANALYZER ASSY, CERT |
| EX800000114 | 1 to 12 | F | 19 Nov 24 | SCS ASSY, ENCLOSURE, J22, CERTIFICATION |
| EX870000015 | 1 to 1 | С | 18 Nov 24 | T-PLATE ASSY, 0.8m |
| EX8700000023 | 1 to 1 | В | 18 Nov 24 | WINDOW, WDG 1.86-2°, Ø25.4x1.5-3.5mm, |
| | | | | GENERIC |
| EX800000092 | 1 to 3 | G | 18 Nov 24 | ANALYZER ASSY, 0.8M, UNIT SEALING |
| EX800000093 | 1 to 4 | E | 18 Nov 24 | ANALYZER SYSTEM ASSY, UNIT SEALING |

