

Klassifizierung nach Druckgeräterichtlinie 2014/68/EU

Wir als Hersteller von Produkten für die Flüssigkeitsanalyse erklären, dass das gelieferte Produkt **Flowfit COA30-**11**4** gemäß Artikel 4, Absatz 3, Druckgeräterichtlinie 2014/68/EU, nach guter Ingenieurspraxis gefertigt wurden und damit nicht CE-kennzeichnungspflichtig ist.


Diese Herstellererklärung ist nur gültig für das oben genannte Produkt.

Bewertungsergebnis

Max. zulässiger Druck:	2 bar
Volumen/ Nennweite:	max. 30 ml/ 3 DN
Mediumeigenschaft nach Art. 13:	Gruppe 1 – gefährlich
Mediumzustand:	gasförmig oder flüssig, wenn PD > 0.5 bar
Qualitätssicherung:	Keine nach Richtlinie 2014/68/EU
Klassifizierung:	Artikel 4, Absatz 3, gute Ingenieurspraxis
Konformitätsbewertungsverfahren:	Kein
Aufgaben des Herstellers:	Keine
Aufgaben der benannten Stelle:	Keine

Diagramm 6, Art. 4, Anhang II

Rohrleitung, Fluid: Gas, Fluidgruppe: 1 (gefährlich)



Endress+Hauser Conducta GmbH+Co. KG
Dieselstraße 24, 70839 Gerlingen, Germany

Datum

Date

i.V. Uwe Rößiger
Manager Certifications & Approvals

Zertifikats-Nr. | Certificate-no.
HE_02055_01.25

Classification according to Pressure Equipment Directive 2014/68/EU

We as a manufacturer of liquid analysis products certify that the provided product **Flowfit COA30-**11**4** were manufactured according to article 4, paragraph 3, Pressure Equipment Directive 2014/68/EU, with sound engineering practice and is not subjected to CE marking.


This manufacturer's declaration is only valid for above mentioned product.

Assessment Result

Max. allowable pressure:	2 bar
Volume/ Diameter:	max. 30 ml/ 3 DN
Fluid group acc. to art. 13:	Group 1 - dangerous
Condition of fluid:	gaseous or liquid, if PD > 0.5 bar
Quality assurance:	None acc. to Directive 2014/68/EU
Classification:	Article 4, paragraph 3, sound engineering practice
Conformity Assessment Procedure:	None
Duties of the manufacturer:	None
Duties of the Notified Body:	None

Table 6, Art. 4, Annex II

Piping, Fluid: gaseous, Fluid group: 1 (dangerous)



2025-04-22

i.V. Marco Rottmann
Certifications & Approvals

Endress+Hauser 
People for Process Automation