

Company **Endress+Hauser SE+Co. KG, Hauptstrasse 1, 79689 Maulburg**

declares as manufacturer, that for the following products

Product **Cerabar**
PMP23

and

Deltabar S

FMD78

the roughness of the process-isolating diaphragm (sensor membrane) $R_a \leq 0.38 \mu\text{m}$ (15.0 μin)

the roughness of other wetted surfaces (excluding weld seam) $R_a \leq 0.76 \mu\text{m}$ (29.9 μin)

This declaration of compliance is exclusively valid for the listed products in delivery status.

Maulburg, 16-Mai-2025
Endress+Hauser SE+Co. KG

i.V. Stefan Jäger
Dept. Manager R&D Qualitymanagement/ FSM

i.V. Lars Kretschmer
Dept. Manager R&D Devices Pressure