

Table of Contents

Scope of Registration 960017312-C.pdf	2
Manufacturing Locations.pdf	4
01_BC_0F18141.51R3_Cerabar-Ceraphant_T.pdf	5
02_AB_0F18141.52R3_Cerabar-Ceraphant_T.pdf	6
03_SK_0F18141.53R3_Cerabar-Ceraphant_T.pdf	8
04_MB_0F18141.54R3_Cerabar-Ceraphant_T.pdf	9
05_ON_0F18141.5R3_Cerabar-Ceraphant_T.pdf	10
06_QC_0F18141.56R3_Cerabar-Ceraphant_T.pdf	11
07_NB_0F18141.57R3_Cerabar-Ceraphant_T.pdf	13
08_NS_0F18141.58R3_Cerabar-Ceraphant_T.pdf	14
09_PEI_0F18141.59R3_Cerabar-Ceraphant_T.pdf	15
10_NL_0F18141.50R3_Cerabar-Ceraphant_T.pdf	16
11_YK_0F18141.5YR3_Cerabar-Ceraphant_T.pdf	17
12_NT_0F18141.5TR3_Cerabar-Ceraphant_T.pdf	18
13_NU_0F18141.5NR3_Cerabar-Ceraphant_T.pdf	19

Doc. #960017312

SCOPE OF CRN REGISTRATION

Product 1:	CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMC11, OPMC11, PMC21, OPMC21
Product 2:	CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B
Code of Construction:	ASME B31.3, ASME B31.1

PRODUCT DESCRIPTION

Design Report Number	Process Connection		Materials of Construction ¹⁾	Maximum Design Pressure (bar) and Temperature (°C) ²⁾	Drawing	Shop level drawing		Type
	Design	Pressure Components						
960017312-01 Revision 2	ASME B1.20.1-2012 NPT Thread	Thread ASME MNPT1/4, bore 3mm (0.12"), 316L	UNS S31603, A-479, (316L, 1.4435) UNS S31603, A-479, (316L, 1.4404)	40 bar -40°C up to 100°C	960016340	960017724	960017721	PMC11, OPMC11, PMC21, OPMC21, PTC31B, OPTC31B
		Thread ASME MNPT1/2, bore 11.4mm (0.45"), 316L				960017726	960017723	
		Thread ASME MNPT1/2 FNPT1/4 (female), 316L				960017725	960017722	
960017312-02 Revision 1	ASME B1.20.1-2012 NPT Thread	Thread ASME MNPT1/4, bore 3mm (0.12"), 316L	UNS S31603, A-479, (316L, 1.4435) UNS S31603, A-479, (316L, 1.4404)	25 bar -40°C up to 100°C	960017295	960017717	960017720, 960017719, 960017714, 960017715	PMP11, OPMP11, PMP21, OPMP21, PTP31B, OPTP31B
		Thread ASME MNPT1/2, bore 11.4mm (0.45"), 316L		100 bar -40°C up to 100°C				
		Thread ASME MNPT1/2 FNPT1/4 (female), 316L		25 bar -40°C up to 100°C				
				100 bar -40°C up to 100°C				
				25 bar -40°C up to 100°C				
				100 bar -40°C up to 100°C				
R-1218A Revision 1	ISO228 EN837 Thread	Thread ISO228 G1/2, EN837, 316L	UNS S31603, A-479, (316L, 1.4435) UNS S31603, A-479, (316L, 1.4404)	25 bar -40°C up to 100°C	961004126	961004124	960017720, 960017719, 960017714, 960017715	
				100 bar -40°C up to 100°C				
960017312-03 Revision 1	ASME B1.20.1-2012 NPT Thread	Thread ASME MNPT1/4, bore 3mm (0.12"), 316L	UNS S31603, A-479, (316L, 1.4435) UNS S31603, A-479, (316L, 1.4404)	400 bar -40°C up to 100°C	960017296	960017485	960017484, 960017504	PMP11, OPMP11, PMP21, OPMP21, PTP31B, OPTP31B
		Thread ASME MNPT1/2, bore 11.4mm (0.45"), 316L				960017489		
		Thread ASME MNPT1/2 FNPT1/4 (female), 316L				960017492		
R-1218B Revision 1	ISO228 EN837 Thread	Thread ISO228 G1/2, EN837, 316L	UNS S31603, A-479, (316L, 1.4435) UNS S31603, A-479, (316L, 1.4404)	400 bar -40°C up to 100°C	961004128	961004127	960017484, 960017504	PMP11, OPMP11, PMP21, OPMP21, PTP31B, OPTP31B
960017312-04 Revision 1	ISO 228 Thread	Thread ISO228 G1 seal conical metal joint, weld-in adapter, 316L	UNS S31603, A-479, (316L, 1.4435) UNS S31603, A-479, (316L, 1.4404)	100 bar -40°C up to 100°C	960017433	960017681	960017708, 960017709, 960017687	PMP23, OPMP23, PMP43, OPMP43, PTP33B, OPTP33B
				25 bar -40°C up to 100°C				
		Thread ISO228 G1 seal O-ring, weld-in adapter, 316L		40 bar -40°C up to 100°C				
		25 bar -40°C up to 100°C						
		100 bar -40°C up to 100°C						
		25 bar -40°C up to 100°C						
DIN13 Thread	Thread M24, 316L, seal O-ring, weld-in adapter			100 bar -40°C up to 100°C		960017684	960017710, 960017711, 960017686	
				25 bar -40°C up to 100°C		960017684	960017710, 960017711, 960017686	
				25 bar -40°C up to 100°C		960017680	960017686	

**THIS IS PART OF CRN
0F18141.5R3**
Technical Standards and Safety Authority
Boilers and Pressure Vessels Safety Program

PRODUCT DESCRIPTION

Design Report Number	Process Connection		Materials of Construction ¹⁾	Maximum Design Pressure (bar) and Temperature (°C) ²⁾	Drawing	Shop level drawing		Type
	Design	Pressure Components						
960017312-05 Revision 1	DIN11851	DIN11851 DN50 (2") PN25, 316L	UNS S31603, A-479, (316L, 1.4435) UNS S31603, A-479, (316L, 1.4404)	25 bar -40°C up to 100°C	960017420	960017650	960017639	PMP23, OPMP23, PMP43, OPMP43, PTP33B, OPTP33B
		DIN11851 DN25 (1") PN40, 316L		40 bar -40°C up to 100°C		960017647	960017628	
		DIN11851 DN40 (1-1/2") PN40, 316L		25 bar -40°C up to 100°C		960017646	960017637	
				40 bar -40°C up to 100°C		960017649	960017629	
				25 bar -40°C up to 100°C		960017648	960017638	
				40 bar -40°C up to 100°C		960017640	960017625	
960017312-06 Revision 1	ISO2825 ³⁾	Tri-Clamp ISO2825 DN18-22 (3/4"), 316L		25 bar -40°C up to 100°C	960017297	960017641	960017634	
		Tri-Clamp ISO2825 DN25-38 (1-1/2"), 316L		40 bar -40°C up to 100°C		960017643	960017626	
		Tri-Clamp ISO2825 DN40-51 (2"), 316L		25 bar -40°C up to 100°C		960017642	960017635	
				40 bar -40°C up to 100°C		960017644	960017627	
				25 bar -40°C up to 100°C		960017645	960017636	
				40 bar -40°C up to 100°C		960017614	960017623	
960017312-07 Revision 1	GEA Tuchenhagen ³⁾	Varivent F pipe DN25-32 PN40, 316L	25 bar -40°C up to 100°C	960017419	960017615	960017630		
		Varivent N pipe DN40-162 PN40, 316L	40 bar -40°C up to 100°C		960017619	960017624		
			25 bar -40°C up to 100°C		960017616	960017631		
			25 bar -40°C up to 100°C		960017655	960017633		
			10 bar -40°C up to 100°C		960017652	960017632, 960017651, 960017653, 960017654		
		960017312-08 Revision 1	SMS ⁴⁾		SMS 1-1/2" PN25, 316L	40 C up to 100°C	960017421	
Endress+Hauser UNI	Universal adapter 44mm 316L, weld-in adapter D65, D85		10 bar -40 C up to 100°C					

THIS IS PART OF CRN 0F18141.5R3
 Technical Standards and Safety Authority
 Boilers and Pressure Vessels Safety Program

- Note 1)** The following additional alloys of superior or equivalent properties to those listed above are also included in the scope of registration as follows:
 N06022, B574/B564, (C-22, 2.4602); N10276, B574/B564, (C-276, 2.4819); S31803, A479/A182, F51, (318LN, 1.4462); S32750, A240/A182, F53, (2507, 1.4410); S32760, A479/A182, F55, (Zeron 100, 1.4501); N06600, B166/B564, (Inconel 600, 2.4816); N06625, B446/B564, (Inconel 625, 2.4856); N08825, B425/B564, (Inconel 825, 2.4858); S30400/S30403, A479/A182, (304/304L, 1.4301/1.4306/1.4307);
- Note 2)** During the Steam-in-Place (SIP) cleaning procedure models PMP23, PMP43 and PTP33B may be subjected to temperatures up to 149 °C (300 °F).
- Note 3)** The assembly clamp is not part of the CRN and is not manufactured by ENDRESS+HAUSER. Pressure-Temperature ratings may be limited by the clamp type used in the joint assembly. Consult clamp Manufacturer for clamp Pressure-Temperature ratings.
- Note 4)** ENDRESS+HAUSER supplies only the SMS nut and female part of the ferrule. The SMS mating ferrule (male part) is not part of the CRN and is not manufactured and not delivered by ENDRESS+HAUSER. Pressure-Temperature ratings may be limited by the SMS mating ferrule used in the joint assembly. Consult mating ferrule manufacturer for SMS mating ferrule Pressure-Temperature ratings.
- Note 5)** See Attached List of Endress+Hauser Manufacturing locations applicable to this CRN.

WORLDWIDE LOCATIONS APPENDIX

**ENDRESS+HAUSER SE+Co. KG MANUFACTURING LOCATIONS
& CERTIFYING AUTHORITIES**

(rev. 04/June/2025)

Endress+Hauser SE+Co. KG

Hauptstrasse 1
79689 Maulburg, Germany
ISO 9001 Certified by SQS

Endress+Hauser (Suzhou) Automation Instrumentation Co. Ltd.

China – Singapore Industrial Park (SIP)
Su Hong Zhong Lu No. 491
Jiangsu Province, 215021 Suzhou, P.R. China
ISO 9001 Certified by SQS

Endress+Hauser (India) Automation Instrumentation Pvt. Ltd.

M-192, Waluj MIDC
Chhatrapati Sambhajinagar – 431136,
Maharashtra State, India
ISO 9001 Certified by SQS

Endress+Hauser Yamanashi Co., Ltd.

862-1 Mitsukunugi
Sakaigawa-cho
Fuefuki-shi
406 0846 Yamanashi, Japan
ISO 9001 Certified by SQS

**Endress+Hauser (USA)
Automation Instrumentation Inc.**

2340 Endress Place
Greenwood, Indiana 46143, USA
ISO 9001 Certified by SQS

**Endress+Hauser (Brazil)
Instrumentação e Automação Ltda.**

Estrada Municipal Antonio Sesti
600 Bairro Recreio Costa Verde
Itatiba/SP – 13254-085, Brazil
ISO 9001 Certified by SQS

Endress+Hauser SE+Co. KG

Queremathe 2
14532 Stahnsdorf
Germany
ISO 9001 Certified by SQS

ATTN: TSSA BPV NATIONAL REGISTRATION -
TECHNICAL STANDARDS & SAFETY AUTHORITY
345 CARLINGVIEW DR
TORONTO ON M9W 6N9

Date: 22-Jan-2026
TSBC Account #: 061440
TSBC Admin Number: 121662
Canadian Registration Number: 0F18141.51

Re: Application for Design Registration

The design, as detailed in your Design Portal application 0F18141.5R3 - ENDRESS+HAUSER SE+Co. KG for a Pressure Fitting is registered with the following notes and considerations:

Registered To:	ENDRESS+HAUSER SE+Co. KG
Project Name:	0F18141.5R3 - ENDRESS+HAUSER SE+Co. KG
Drawing #:	Scope of CRN Registration
Drawing Revision:	C

Conditions of Registration:

(1) Fitting Registration Expiry Date: 27-Nov-2035 (2) The registration is valid until the indicated expiry date only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date. Should the approval of quality management system lapse before the expiry date indicated above, this registration shall become void.

Reviewer's Notes:

Any additional conditions and considerations from the initial province of registration shall apply to this BC registration.

Full details of this submission including the scope of registration, design conditions, fabrication details, and calculations pertaining to this design are located in the above Admin Number on the Design Portal. For all other enquiries, please contact eim@technicalsaftybc.ca.

The Engineering Information Management Team

February 26, 2026

Attention: Tanya Francis
TECHNICAL STANDARDS & SAFETY AUTHORITY
345 CARLINGVIEW DRIVE
TORONTO, ON M9W 6N9

The design submission, Tracking Number 2025-08186, Web Portal Number 2025-S6621, originally received on December 19, 2025 was surveyed and accepted for registration as follows:

CRN : 0F18141.52 **Accepted on:** February 26, 2026
Reg Type: RENEWAL **Expiry Date:** November 28, 2035
Document No. 960017312 (Scope of CRN Registration) Rev C As Noted
Fitting type: Cerabar, Ceraphant Pressure Sensors

Design registered in the name of : ENDRESS + HAUSER SE + CO KG

The registration is conditional on your compliance with the following notes:

- *This is a renewal of CRN 0F18141.52, including a material grade change to 316L and addition of ASME B31.1 Code of Construction.*

- *Per your confirmation, it is my understanding that during the Steam-in-Place (SIP) cleaning procedure models PMP23, PMP43, and PTP33B may be subjected to pressure not exceeding 4 barg (58 psig) maximum and temperatures up to 149°C (300°F).*

- *Material grades ASTM A479/A182 F55 and ASTM B166/B564 (Inconel 600, 2.4816) N06625 are unlisted in ASME B31.1, 2024 edition, and shall not be used for B31.1 construction as they have not been qualified in accordance with paragraph 123.1.2.*

As indicated on AB-41 Statutory Declaration or AB-351 Declaration of Conformity form and submitted documentation, the code of construction are ASME B31.3 and ASME B31.1.

- *It is our understanding that the fitting(s), included as the scope of this submission, that is(are) subject to the Safety Codes Act shall comply with the requirements of the indicated Standard or Code of Construction on the AB-41 Statutory Declaration or AB-351 Declaration of Conformity as supported by the attached data which identifies the dimensions, materials of construction, press./temp. ratings and the basis for such ratings, and the identification marking of the fittings.*

- *This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration or AB-351 Declaration of Conformity form.*

- *This registration is valid only until the indicated expiry date and only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency, and maintains a valid Certification of Authorization Permit if required by the jurisdiction where manufacturing takes place, until that date.*

- *Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.*

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

If you have any questions don't hesitate to contact me by phone at (587) 686-9387 or fax (780) 437-7787 or e-mail Stolarchuk@absa.ca.

February 26, 2026

Sincerely,



STOLARCHUK, ALEX
DOP Cert. No. D00011392

REGISTRATION OF A PRESSURE FITTING DESIGN

January 9, 2026

TSSA
345 Carlingview Dr.
Toronto, ON
Canada
M9W 6N9

Attention: Tanya Francis

File Number: 104168

Re: Manufacturer: Endress+Hauser SE+Co. KG
Item: CERABAR / CERAPHANT Sensors
Catalog or Drawing: Per Scope of CRN Registration Rev. C (18/Sept/2025) & Design Reports

TSASK Codes and Standards Compliance has registered the design listed above in accordance with The Boiler and Pressure Vessel Act and Regulations and CSA B51. The Canadian Registration Number (CRN) is:

OF18141.53

Expiry Date: 2035-11-28

Please note that every fitting shall be constructed in strict accordance with the registered design.

Fitting registrations are required to be resubmitted for validation after ten (10) years from the registration date in accordance with CSA B51, Clause 4.2.1.

Should you require anything further, please do not hesitate to contact the Codes and Standards Compliance Office at your convenience.

Yours truly,



Athan Syrgiannis, P.Eng.

Codes and Standards Compliance

Remarks:

CRN renewal.

CRN issuance based upon registration by another province under the interprovincial agreement. Conditional upon compliance with the notes on the TSSA registration.

See Worldwide Locations Appendix (rev. 04/June/2025) for list of manufacturing locations.



January 27, 2026

TSSA
345 Carlingview Drive
Toronto, ON M9W 6N9

Dear Tanya Francis,

Re: Reciprocal CRN Registration in Manitoba

As indicated by the Regulatory Reconciliation and Cooperation Table and the Reconciliation Agreement for the Canadian Registration Number (CRN) for Pressure Equipment, the design reviews conducted and accepted by the Canadian province or territory, or their delegated safety authority, will be mutually recognized in the Province of Manitoba. If a registration is conditionally based on compliance with the notes set by the original issuing Jurisdiction, such compliance shall be applied the same to this Province.

Your submission has been registered, as follows:

File Number:	74-R5796
CRN:	0F18141.54
Scope:	Scope of CRN Registration
Manufacturer:	ENDRESS+HAUSER SE+Co. KG
Expiry Date:	28 November 2035

Along with this letter is the invoice for registration.

In addition, every Pressure Vessel, Boiler, and Heat Exchanger shall be stamped with the registration number and as required by CSA Code B51, a Manufacturer's Data Report (MDR) must be forwarded to this office immediately at the time a unit is shipped to Manitoba. Send your MDR to gasupport@gov.mb.ca. In your subject line, indicate "*Manufacturer's Data Report-CRN No.*" A fee shall be billed to the Manufacturer to process data reports in accordance with the Steam and Pressure Plants Regulation section 17.1.

Please contact gasupport@gov.mb.ca for any questions or concerns.

Inspection and Technical Services

Labour and Immigration
508 – 401 York Avenue, Winnipeg, MB R3C 0P8
T (204) 945-3373 | F (204) 948-2089



345 Carlingview Drive
Toronto, Ontario M9W 6N9
Tel.: 416.734.3300
Fax: 416.231.1626
Toll Free: 1.877.682.8772

www.tssa.org

November 28, 2025

ROUND ENGINEERING INC
10 SEGWUN RD
WATERDOWN ON L8B 0K6

Workorder Type: Registration - Fitting(Conventional)
Workorder No: 14817163
Your Reference No.: R-2250
Registered to: ENDRESS+HAUSER SE+Co. KG

Dear SCOTT ISLIP,

Technical Standards and Safety Authority (TSSA) is pleased to inform you that your submission has been reviewed and registered as follows:

CRN : 0F18141.5R3

Main Design No.: Scope of CRN Registration

Expiry Date: Nov 28, 2035

Please be advised that a valid quality control system must be maintained for the fitting registration to remain valid until the expiry date.

- Renewal of CRN 0F18141.5, including a material grade change to 316L and addition of ASME B31.1 to the registration

The stamped copy of the approved registration and the invoice are mailed separately (There will be no hard copies for electronic submissions). Should you have any questions or require further assistance, please contact a Customer Service Advisor at 1.877.682.TSSA (8772) or e-mail customerservices@tssa.org. We will be happy to assist you. When contacting TSSA regarding this file, please refer to the Service Request number provided above.

Yours truly,

Hammad Wasif , M.Eng., P.Eng.
Engineer, BPV
Tel. : +1 416-734-3441
Email : hwasif@tssa.org

Montréal, le 6 février 2026.

TANYA FRANCIS
TECHNICAL STANDARDS & SAFETY AUTHORITY
345 CARLINGVIEW DRIVE
TORONTO ON
CANADA M9W 6N9

Manufacturer : ENDRESS + HAUSER GMBH + CO KG
1 HAUPTSTRASSE
MAULBURG GERMANY
79689

OUR REFERENCE : 940896
Design number : Scope of CRN Registration
0F18141.5R3

Subject: Design registration confirmation

Hi,

We wish to inform you that your design registration application has been evaluated and that it was registered under the following Canadian Registration Number (CRN): **0F18141.56**.

The following is a reminder of your obligations regarding certain requirements of the regulation respecting pressure vessels, and the referenced codes and standards:

- The manufacturer must maintain a valid quality control program to manufacture equipment according to the CRN.
- The CRN remains valid as long as there are no changes to the design calculations that might affect the pressure boundary. The design registration of fittings expires 10 years after acceptance. It must, therefore, be resubmitted for validation.
- The manufacturer shall submit a copy of the *Manufacturer's Data Report* to us for each boiler or pressure vessel manufactured according to this CRN within 30 days following the signing of this report.
- The drawing number and the revision number registered under this CRN must be indicated on the *Manufacturer's Data Report* for equipment manufactured according to the CRN.

This notice of approval does not relieve the manufacturer of their responsibilities with respect to the design or fabrication of equipment manufactured according to this CRN.

Yours sincerely,

Direction des équipements sous pression

Montréal

255, boul. Crémazie Est, 2ième étage
Montréal (Québec) H2M 1L5
Téléphone : 514 873-2546
Sans frais : 1 866 262-2084
enregistrementdesplans@rbq.gouv.qc.ca
www.rbq.gouv.qc.ca

Montréal, 6 février 2026.

TANYA FRANCIS
TECHNICAL STANDARDS & SAFETY AUTHORITY
345 CARLINGVIEW DRIVE
TORONTO ON
CANADA M9W 6N9

Fabricant : ENDRESS + HAUSER GMBH + CO KG
1 HAUPTSTRASSE
MAULBURG GERMANY
79689

Numéro de dossier : 940896
Numéro(s) de dessin(s) : Scope of CRN Registration
0F18141.5R3

Objet : Enregistrement des plans et devis – Confirmation de l'enregistrement

Bonjour,

Nous vous informons que votre demande d'enregistrement de plans et devis a été traitée et que cette conception a été enregistrée sous le numéro d'enregistrement canadien (NEC\CRN) suivant : **0F18141.56**.

Nous portons votre attention sur certaines exigences réglementaires concernant les installations sous pression, ainsi que des codes et normes qui y sont associés :

- Le fabricant doit maintenir un programme de contrôle de la qualité valide pour fabriquer un équipement selon ce NEC;
- Ce numéro d'enregistrement demeure valide tant et aussi longtemps que les paramètres de conception demeurent inchangés. Dans le cas d'accessoires, l'enregistrement est valide pour une durée de 10 ans à partir de la date d'enregistrement. Les documents de conception doivent alors être resoumis pour validation;
- Le fabricant doit nous transmettre une copie de la *Déclaration de conformité du constructeur (Manufacturer's Data Report)* pour chaque appareil ou chaudière fabriqué selon ce NEC dans les 30 jours suivant la signature de cette déclaration;
- Le numéro de dessin enregistré et le numéro de révision doivent être indiqués sur la déclaration de conformité pour les équipements fabriqués selon ce NEC.

Le présent avis d'approbation ne dégage pas le fabricant de ses responsabilités quant à la conception ou à la construction des équipements ou d'accessoires fabriqués selon un NEC.

Salutations distinguées,

Direction des équipements sous pression

Montréal

255, boul. Crémazie Est, 2^{ième} étage

Montréal (Québec) H2M 1L5

Téléphone : 514 873-2546

Sans frais : 1 866 262-2084

enregistrementdesplans@rbq.gouv.qc.ca

www.rbq.gouv.qc.ca

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

New Brunswick X Nova Scotia X Prince Edward Island X Newfoundland and Labrador X
 Nunavut X Yukon X Northwest Territories X

Manufacturers Name: Endress+Hauser SE+Co. KG, See also "Worldwide Locations Appendix"	
Manufacturers Address: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
Plant Locations: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
<p align="center">Category of Fittings to be registered. Circle one Category only</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p align="center">Title of the Standard of Construction</p> <p>ASME B31.3 <i>ASME B31.1</i> </p>
Show Manufacturers Name, Trademark, or Logo as it will appear on the product	
<p>Endress+Hauser </p> <p>People for Process Automation</p>	
Type of Construction	
Forged <input type="checkbox"/> Welded <input checked="" type="checkbox"/> Wrought <input checked="" type="checkbox"/> Cast <input type="checkbox"/> Other <input type="checkbox"/> Describe other:	
List of supporting documentation and identification of the actual items to be registered:	
Cerberar, Ceraphant - Scope of Registration, Reports, Drawings, Calculations	

Declaration:

I **Florian Palatini** (see note 3) employed by **Endress+Hauser SE+Co. KG** and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by **SQS ISO 9001:2015** as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: _____
 Declared before me at _____
 This _____ day of _____ AD _____
 Commissioner of Oaths _____
 Or Notary Public: (sign) _____
 (Affix Official seal to the right)

Use this space for the Official Seal

New Brunswick
 Nouveau Brunswick

DEPT OF JUSTICE PUBLIC SAFETY
BOILER & PRESSURE VESSEL ACT

REGISTRATION ONLY

CRN **0F18141.57 Rev.3**

CHIEF BOILER INSPECTOR
 DATE **1/2/26**

Sect-1.0 - Fittings Rev.2

BLRs PVs

FITTINGS NUCLEAR COMPONENTS

This space for Regulatory Authority Use
 This registration must be revalidated after ten (10) years from the date of acceptance.

CRN: **0F18141.5 Rev.3**

FID#: _____


Notes:

- All Fittings shall be registered in the name of the Manufacturer.
- Each Category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
- The Declaration shall be made by the person having full authority and responsibility for the quality of the end product.
- Quality Control programs shall be resubmitted for validation.

Scope of registration 960017312 rev.C, CRN renewal.
 Product 1: CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMC11, OPMC11, PMC21, OPMC21. Product 2: CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B. 7 plant locations (DGallant)

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

New Brunswick x Nova Scotia x Prince Edward Island x Newfoundland and Labrador x
 Nunavut x Yukon x Northwest Territories x

Manufacturers Name: <u>Endress+Hauser SE+Co. KG, See also "Worldwide Locations Appendix"</u>	
Manufacturers Address: <u>Hauptstrasse 1 79889 Maulburg, Germany, See also "Worldwide Locations Appendix"</u>	
Plant Locations: <u>Hauptstrasse 1 79889 Maulburg, Germany, See also "Worldwide Locations Appendix"</u>	
<p align="center">Category of Fittings to be registered. Circle one Category only</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories</p> <p>N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> , (Meeting CNSC or ASME requirements)</p>	<p align="center">Title of the Standard of Construction</p> <p>ASME B31.3 ASME G31.1 <i>[Signature]</i></p>
<p>Show Manufacturers Name, Trademark, or Logo as it will appear on the product</p> <p align="center">Endress+Hauser  People for Process Automation</p>	<p align="center">Type of Construction</p> <p>Forged <input type="checkbox"/> Welded <input type="checkbox"/> Wrought <input type="checkbox"/> Cast <input type="checkbox"/> Other <input type="checkbox"/> Describe other:</p>
<p>List of supporting documentation and identification of the actual items to be registered:</p> <p>Cerabar, Ceraphant - Scope of Registration, Reports, Drawings, Calculations</p>	

Declaration:
 I Florian Pelatin (see note 3) employed by Endress+Hauser SE+Co. KG and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by SQS ISO 9001:2015 as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: [Signature] 04. Juni 2025, Schopfheim

Declared before me at _____

This _____ day of _____ AD _____

Use this space for the Official Seal

Commissioner of Oaths


Or Notary Public: (sign) _____

(Affix Official seal to the right)

<p>CRN: <u>OF18141.5 Rev.3</u></p> <p>FID#: _____</p> <p>Notes:</p> <ol style="list-style-type: none"> All Fittings shall be registered in the name of the Manufacturer. Each Category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. The Declaration shall be made by the person having full authority and responsibility for the quality of the end product. Quality Control programs shall be resubmitted for validation. <p>Scope of registration 950017312 rev.C, CRN renewal. Product 1: CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMG11, OPMP11, PMP21, OPMP21. Product 2: CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B. 7 plant locations (DGallant)</p>	<p align="center">This space for Regulatory Authority use. This registration must be revalidated after ten (10) years from the date of completion.</p> <p align="center">NOVA SCOTIA BOILER & PRESSURE VESSEL Date <u>Jan. 2/26</u> C.R.N. <u>OF18141.5B Rev. 3</u> Dwg. <u>as described</u> Signed <u>[Signature]</u> 1 of 1 Part Sect. 19 Fittings Rev.2</p>
---	--

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

New Brunswick X Nova Scotia X Prince Edward Island X Newfoundland and Labrador X
 Nunavut X Yukon X Northwest Territories X

Manufacturers Name: <u>Endress+Hauser SE+Co. KG</u> , See also "Worldwide Locations Appendix" Manufacturers Address: <u>Hauptstrasse 1 78689 Maulburg, Germany</u> , See also "Worldwide Locations Appendix" Plant Locations: <u>Hauptstrasse 1 79689 Maulburg, Germany</u> , See also "Worldwide Locations Appendix"	
Category of Fittings to be registered. Circle one Category only A. Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B. Flanges; all flanges C. Valves; all line valves D. Expansion joints, flexible connections, and hose assemblies; all types E. Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F. Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G. Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H. Pressure retaining components that do not fall into one of the above categories N. Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)	Title of the Standard of Construction ASME B31.3 ASME B31.1 <i>i.v. [Signature]</i>
Show Manufacturers Name, Trademark, or Logo as it will appear on the product <div style="text-align: center;"> Endress+Hauser  People for Process Automation </div>	Type of Construction Forged <input type="checkbox"/> Welded <input type="checkbox"/> Wrought <input type="checkbox"/> Cast <input type="checkbox"/> Other <input type="checkbox"/> Describe other:
List of supporting documentation and identification of the actual items to be registered: Cerabar, Ceraphant - Scope of Registration, Reports, Drawings, Calculations	


Declaration:
 I Florian Palatini (see note 3) employed by Endress+Hauser SE+Co. KG and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by ISO 9001:2015 as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: [Signature] 04. Juni 2025, Schopthheim
 Declared before me at _____
 This _____ day of _____ AD _____
 Commissioner of Oaths
 Or Notary Public: (sign) _____
 (Affix Official seal to the right)

This space for Regulatory Authority use. This registration must be revalidated after ten (10) years from the date of acceptance. CRN: <u>0F18141.5 Rev.3</u> FID#: _____ Notes: 1. All Fittings shall be registered in the name of the Manufacturer. 2. Each Category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. 3. The Declaration shall be made by the person having full authority and responsibility for the quality of the end product. 4. Quality Control programs shall be resubmitted for validation. Scope of registration: 960017312 rev.C, CRN renewal. Product 1: CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMC11, OPMP11, PMC21, OPMP21. Product 2: CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B. 7 plant locations (DGallant)	ACCEPTED PROVINCE OF PRINCE EDWARD ISLAND C.R.N.: <u>0F18141.57R03</u> DATE: <u>JAN 27/25</u> SIGNED: <u>[Signature]</u> INSPECTION SERVICES SECTION BOILER/PRESSURE VESSEL DEPARTMENT
---	--

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

New Brunswick x Nova Scotia x Prince Edward Island x Newfoundland and Labrador x
 Nunavut x Yukon x Northwest Territories x

Manufacturers Name: Endress+Hauser SE+Co. KG, See also "Worldwide Locations Appendix"	
Manufacturers Address: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
Plant Locations: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
<p align="center">Category of Fittings to be registered. Circle one Category only</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p align="center">Title of the Standard of Construction</p> <p>ASME B31.3 ASME B31.1 <i>[Signature]</i></p>
<p align="center">Show Manufacturers Name, Trademark, or Logo as it will appear on the product</p> <p align="center">Endress+Hauser  People for Process Automation</p>	<p align="center">Type of Construction</p> <p>Forged <input type="checkbox"/> Welded <input type="checkbox"/> Wrought <input type="checkbox"/> Cast <input type="checkbox"/> Other <input type="checkbox"/> Describe other:</p>
<p align="center">List of supporting documentation and identification of the actual items to be registered:</p> <p>Cerabar, Ceraphant - Scope of Registration, Reports, Drawings, Calculations</p>	

Declaration:
 I Florian Palatni (see note 3) employed by Endress+Hauser SE+Co. KG and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by ISO 9001:2015 as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.


Signature of Declarer: [Signature] 04. Juni 2025, Schopfheim

Declared before me at _____
 This _____ day of _____ AD _____
 Commissioner of Oaths
 Or Notary Public: (sign) _____
 (Affix Official seal to the right)

<p align="center">This space for Regulatory Authority of Newfoundland and Labrador This registration must be revalidated after ten (10) years from the date of acceptance.</p>	
CRN: <u>OF18141.5 Rev.3</u> FID#: _____ Notes: 1. All Fittings shall be registered in the name of the Manufacturer. 2. Each Category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. 3. The Declaration shall be made by the person having full authority and responsibility for the quality of the end product. 4. Quality Control programs shall be resubmitted for validation. Scope of registration 960017312 rev.C, CRN renewal, Product 1: CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMC11, OPMP11, PMC21, OPMP21, Product 2: CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B, 7 plant locations (DGallant)	<p align="center">Registered <u>OF18141.5 Rev 3</u> Date <u>Jan 6, 2026</u> Engineering and Inspection Services Registered by <u>[Signature]</u> UNDER THE AUTHORITY OF THE PUBLIC SAFETY ACT AND THE BOILER, PRESSURE VESSEL AND COMPRESSED GAS REGULATIONS Sect 1.0 - Fittings Rev.2</p>

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

New Brunswick x Nova Scotia x Prince Edward Island x Newfoundland and Labrador x
 Nunavut x Yukon x Northwest Territories x

Manufacturers Name: Endress+Hauser SE+Co. KG, See also "Worldwide Locations Appendix"	
Manufacturers Address: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
Plant Locations: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
<p align="center">Category of Fittings to be registered. Circle one Category only</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p align="center">Title of the Standard of Construction</p> <p>ASME B31.3 ASME B31.1 <i>[Signature]</i></p>
<p align="center">Show Manufacturers Name, Trademark, or Logo as it will appear on the product</p> <p align="center">Endress+Hauser  People for Process Automation</p>	<p align="center">Type of Construction</p> <p>Forged <input type="checkbox"/> Welded <input checked="" type="checkbox"/> Wrought <input type="checkbox"/> Cast <input type="checkbox"/> Other <input type="checkbox"/> Describe other:</p>
<p>List of supporting documentation and identification of the actual items to be registered:</p> <p>Cerabar, Ceraphant - Scope of Registration, Reports, Drawings, Calculations</p>	

Declaration:
 I Florian Paolini (see note 3) employed by Endress+Hauser SE+Co. KG and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by SQS ISO 9001:2015 as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.


Signature of Declarer: [Signature] 04. Juni 2025, Schopfheim

Declared before me at _____
 This _____ day of _____ AD _____

Use this space for the Official Seal



Commissioner of Oaths
 Or Notary Public: (sign) _____

(Affix Official seal to the right)


<p align="center">This space for Regulatory Authority use. This registration must be revalidated after ten (10) years from the date of acceptance.</p> <p>CRN: <u>0F18141.5 Rev.3</u></p> <p>FID#: _____</p> <p>Notes:</p> <ol style="list-style-type: none"> All Fittings shall be registered in the name of the Manufacturer. Each Category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. The Declaration shall be made by the person having full authority and responsibility for the quality of the end product. Quality Control programs shall be resubmitted for validation. Scope of registration 960017312 rev.C, CRN renewal. Product 1: CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMC11, OPMC11, PMC21, OPMC21. Product 2: CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B. 7 plant locations (DGallant) 	
---	--


UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

New Brunswick x Nova Scotia x Prince Edward Island x Newfoundland and Labrador x
 Nunavut x Yukon x Northwest Territories x

Manufacturers Name: <u>Endress+Hauser SE+Co. KG, See also "Worldwide Locations Appendix"</u>	
Manufacturers Address: <u>Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"</u>	
Plant Locations: <u>Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"</u>	
<p align="center">Category of Fittings to be registered. Circle one Category only</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p align="center">Title of the Standard of Construction</p> <p>ASME B31.3 <i>ASME B31.1</i> </p>
<p align="center">Show Manufacturers Name, Trademark, or Logo as it will appear on the product</p> <p align="center">Endress+Hauser  People for Process Automation</p>	<p align="center">Type of Construction</p> <p>Forged <input type="checkbox"/> Welded <input type="checkbox"/> Wrought <input type="checkbox"/> Cast <input type="checkbox"/> Other <input type="checkbox"/> Describe other:</p>
<p>List of supporting documentation and identification of the actual items to be registered: Cerabar, Ceraphant - Scope of Registration, Reports, Drawings, Calculations</p>	


Declaration:
 I Florian Palatini (see note 3) employed by Endress+Hauser SE+Co. KG and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by SQS ISO 9001:2015 as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer:  04. Juni 2025, Schopfheim
 Declared before me at _____
 This _____ day of _____ AD _____ Use this space for the Official Seal
 Commissioner of Oaths
 Or Notary Public: (sign) _____
 (Affix Official seal to the right)

<p align="center">This space for Regulatory Authority use.</p> <p align="center">This registration must be revalidated after ten (10) years from the date of acceptance.</p> <p>CRN: <u>0F18141.5 Rev.3</u></p> <p>FID#: _____</p> <p>Notes:</p> <ol style="list-style-type: none"> All Fittings shall be registered in the name of the Manufacturer. Each Category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. The Declaration shall be made by the person having full authority and responsibility for the quality of the end product. Quality Control programs shall be resubmitted for validation. <p>Scope of registration 960017312 rev.C, CRN renewal. Product 1: CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMC11, OPMC11, PMC21, OPMC21. Product 2: CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B. 7 plant locations (DGallant)</p>	 Northwest Territories REGISTERED UNDER THE AUTHORITY OF THE BOILER AND PRESSURE VESSEL ACT. C.R.N. <u>0F18141.5 Rev.3</u> SIGNED <u>Matthias Mailman</u> Sect 1.0 - Fittings Rev.2 DATE <u>17/2/26</u>
--	---

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

New Brunswick x Nova Scotia x Prince Edward Island x Newfoundland and Labrador x
 Nunavut x Yukon x Northwest Territories x

Manufacturers Name: Endress+Hauser SE+Co. KG, See also "Worldwide Locations Appendix"	
Manufacturers Address: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
Plant Locations: Hauptstrasse 1 79689 Maulburg, Germany, See also "Worldwide Locations Appendix"	
<p align="center">Category of Fittings to be registered. Circle one Category only</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p align="center">Title of the Standard of Construction</p> <p>ASME B31.3 ASME B31.1 <i>[Signature]</i></p>
<p>Show Manufacturers Name, Trademark, or Logo as it will appear on the product</p> <p align="center">Endress+Hauser  People for Process Automation</p>	<p align="center">Type of Construction</p> <p>Forged <input type="checkbox"/> Welded <input checked="" type="checkbox"/> Wrought <input type="checkbox"/> Cast <input type="checkbox"/> Other <input type="checkbox"/> Describe other:</p>
<p>List of supporting documentation and identification of the actual items to be registered:</p> <p>Cerabar, Ceraphant - Scope of Registration, Reports, Drawings, Calculations</p>	

Declaration:

I Florian Palatini (see note 3) employed by Endress+Hauser SE+Co. KG and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by ISO 9001:2015 as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: [Signature]
 Declared before me at _____
 This _____ day of _____ AD _____
 Commissioner of Oaths
 Or Notary Public: (sign) _____
 (Affix Official seal to the right)

04. Juni 2025, Schopfheim

Use this space for the Official Seal

<p>CRN: <u>0F18141.5 Rev.3</u></p> <p>FID#: _____</p> <p>Notes:</p> <ol style="list-style-type: none"> All Fittings shall be registered in the name of the Manufacturer. Each Category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. The Declaration shall be made by the person having full authority and responsibility for the quality of the end product. Quality Control programs shall be resubmitted for validation. <p>Scope of registration 960017312 rev.C, CRN renewal. Product 1: CERABAR Types: PMP11, OPMP11, PMP21, OPMP21, PMP23, OPMP23, PMP43, OPMP43, PMC11, OPMC11, PMC21, OPMC21. Product 2: CERAPHANT Types: PTP31B, OPTP31B, PTP33B, OPTP33B, PTC31B, OPTC31B. 7 plant locations (DGallant)</p>	<p align="center">This space for Regulatory Authority use. This registration must be revalidated after 5 years from the date of issue.</p> <p align="center">NUNAVUT Boilers and Pressure Vessels Act</p> <p align="center">REGISTERED</p> <p>CRN <u>0F18141.5N Rev.3</u></p> <p>Date <u>01/02/2026</u></p> <p>Signed <u>[Signature]</u> Chief Inspector Sect 1.0 - Fittings Rev.2</p> <p align="center">Territorial Registration Fee</p>
---	--