



防爆合格证

证 号: GYB26.1701U

制 造 商 恩德斯豪斯温度仪表(苏州)有限公司

(地址: 江苏省苏州市苏州工业园区 江田里路 31 号)

产 品 名 称 温度计表头

型 号 规 格 TA30 系列

Ex db IIC Gb (TA30H)

防 爆 标 志 Ex tb IIIC Db (TA30A/TA30D/TA30H)

Ex ia IIIC Da (TA30A/TA30D/TA30H)

产 品 标 准 /

图 样 编 号 10000008487, 10000003875

经图样及技术文件的审查和样品检验, 确认上述产品符合下列标准:

GB/T 3836.1-2021, GB/T 3836.2-2021, GB/T 3836.4-2021, GB/T 3836.31-2021

特颁发此证。

本证书有效期: 2026 年 06 月 15 日 至 2031 年 06 月 14 日

备 注

1. 型号规格说明见本证书附件。
2. 证书编号后缀“U”表明本证书具有限制条件, 详见本证书附件。
3. 产品使用注意事项见本证书附件。
4. 电气参数见本证书附件。

批 准

上海仪器仪表自控系统检验测试所有限公司
国家级仪器仪表防爆安全监督检验站

颁 发 日 期 二 〇 二 六 年 六 月 十 五 日

本证书仅对与认可文件和样品一致的产品有效。



SITIAs
Worldwide Access

EXPLOSION PROTECTION

CERTIFICATE OF CONFORMITY

Cert No. GYB26.1701U

Manufacturer	Endress+Hauser Wetzer(Suzhou)Co.,Ltd. (Address:No.31 Jiang-Tian-Li-Lu,Suzhou Industrial Park,Suzhou, Jiangsu Province)
Product	Connection heads
Model	TA30 series
Ex marking	Ex db IIC Gb (TA30H), Ex tb IIIC Db (TA30A,TA30D and TA30H), Ex ia IIIC Da (TA30A,TA30D and TA30H)
Product standard	/
Drawing number	10000008487, 10000003875

The product was found to comply with the following standard(s):

GB/T 3836.1-2021,GB/T 3836.2-2021,GB/T 3836.4-2021,GB/T 3836.31-2021

Valid until: 2031.06.14

Remarks

- 1.The model designation is specified in the attachment(s) to this certificate.
- 2.The symbol "U" placed after the certificate number denotes schedule of limitations, which are specified in the attachment to this certificate.
- 3.The conditions for use are specified in the attachment(s) to this certificate.
- 4.The electrical parameters are specified in the attachment(s) to this certificate.

Approval


**Shanghai Inspection and Testing Institute of
Instruments and Automation Systems Co., Ltd.**
**National Supervision and Inspection Center for
Explosion Protection and Safety of Instrumentation**
Date of issue 2026.06.15

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.