

Company **Endress+Hauser SE+Co. KG, Hauptstrasse 1, 79689 Maulburg**

being the manufacturer, declares that the “process-contact parts“ of the product

Product **One rod probe**
11371

Material	Order codes	Process-contact parts	FDA conformity
316L	020: 1, 4	Process connection	Not applicable
316Ti	020: 2 030: 1, 2, 3	Process connection	Not applicable
PFA	030: 1, 2, 3	Insulation of rod probe	21 CFR 175.300
Silicone VMQ80	/	Gasket in process connection	21 CFR 177.2600

Regulations are suitable for direct contact with foodstuff and compliant with the FDA Code of Federal Regulations.

Additional information:

The product must be cleaned before use, as appropriate for the application. This declaration of conformity emphasizes that the customer is obliged to test the product regarding its suitability in the application. This declaration of conformity is exclusively valid for the listed products as supplied to customer and without any further modification. Since the conditions of use are beyond Endress+Hauser’s control, Endress+Hauser makes no warranties, expressed or implied, and assumes no liability in connection with any use of this declaration. This declaration of compliance refers to the compliance of the mentioned materials with the listed regulations. No further suitability of the products for a specific purpose shall be derived from this.

Maulburg, 2026-03-16
Endress+Hauser SE+Co. KG

i.V. Thorsten Springmann
Dept. Manager R&D Devices Level Limit

i.V. Stefan Jäger
Dept. Manager R&D Quality Management/FSM