

Certificate

PROFIBUS Nutzerorganisation e.V. grants to

Endress+Hauser Conducta GmbH & Co. KG

Dieselstrasse 24, 70839 Gerlingen, Germany

the Certificate No: **Z01671** for the PROFIBUS Device:

Model Name: Liquistation CSFxx
 Revision: Profile 3.02; SW/FW: 01.16; HW: 01.00
 GSD: EH01155C.GSD, File Version: 20.05.2022
 PA039760.GSD

This certificate confirms that the product has successfully passed the certification tests with the following scope:

| | |
|--|---------------------------------------|
| <input checked="" type="checkbox"/> DP-V0 | MS0, Freeze, Auto_Baud, Set_Slave_Add |
| <input checked="" type="checkbox"/> DP-V1 | MS1, Prm_Block_Structure, MS2, I&M |
| <input checked="" type="checkbox"/> Profile | PROFIBUS PA 3.02 |
| <input checked="" type="checkbox"/> Physical Layer | RS485 |

Test Report Number: 342-1 plus Manufacturer Declaration
 Authorized Test Laboratory: ifak, Magdeburg, Germany

The tests were executed in accordance with the following documents:
 "Test Specifications for PROFIBUS DP Slaves, Version 3.0.9, September 2008" and
 "Test Specifications for PROFIBUS PA Profile 3.0.2, Version 5.0.4, February 2010".

This certificate is granted according to the document:
 "Framework for testing and certification of PROFIBUS and PROFINET products".

For all products that are placed in circulation by **January 24, 2029** the certificate is valid for life.

Karlsruhe, February 24, 2026



(Official in Charge)

Board of PROFIBUS Nutzerorganisation e. V.



(Xaver Schmidt)




(Frank Moritz)