

Installation Instructions Kits for Dipfit CYA10

Spare parts kit for the immersion assembly CYA10



Table of contents


1	Overview	3
2	Intended use	5
3	Authorized installation personnel	5
4	Safety instructions	5
5	Additional documentation	6
6	Disposal	6

1 Overview

1.1 Spare parts kits

Order code	Designation	Scope of delivery
71671555	CYA10 Sealings EPDM →  1,  4, items 3-5 and 7	 A0061103
71671538	CYA10 Sealing oval flange →  1,  4, item 9	 A0061104
71671557	CYA10 Cable gland →  1,  4, item 1	 A0061105
71671577	CYA10 Multifunctional clamp ring →  1,  4, item 8	 A0061106
71671585	CYA10 Hose EPDM, length 1 m →  1,  4, item 2	 A0061107
71717049	CYA10 Hose EPDM, length 3 m →  1,  4, item 2	
71717051	CYA10 Hose EPDM, length 5 m →  1,  4, item 2	
71717052	CYA10 Hose EPDM, length 10 m →  1,  4, item 2	
71671587	CYA10 Protective cap →  1,  4, item 6	 A0061108

1.2 Overview of the kits for the immersion assembly



A0061102

1 CYA10 design

- 1 *Cable gland*
- 2 *Hose*
- 3 *O-ring ID 18.72 W 2.62 OD 23.96*
- 4 *O-Ring ID 25.07 W 2.62 OD 30.31*
- 5 *O-Ring ID 11.00 W 2.00 AD 15.00*
- 6 *Protective cap*
- 7 *O-ring ID 32.99 W 2.62 OD 38.23*
- 8 *Multifunctional clamp ring*
- 9 *Sealing oval flange*

2 Intended use

- The parts in the kits are designed exclusively as maintenance parts for CYA10 submersible assemblies. Any other use is not permitted!
- Only use original parts from Endress+Hauser.
- In the Device Viewer, check if the spare part is suitable for the device in question.

3 Authorized installation personnel

- Installation, commissioning, operation and maintenance of the measuring system may be carried out only by specially trained technical personnel.
- The technical personnel must be authorized by the plant operator to perform the stated tasks.
- The electrical connection may only be established by an electrical technician.
- The technical personnel must have read and understood these Installation Instructions and must follow the instructions they contain.
- Measuring point faults may be repaired only by authorized and specially trained personnel.
- In the case of Ex-certified devices, the technical staff must also be trained in explosion protection.

 Repairs not described in the Operating Instructions provided must only be carried out directly at the manufacturer's site or by the service organization.

4 Safety instructions

WARNING

Risk of injury from high pressure, high temperatures, chemical hazards, or escaping medium!

- Before any maintenance task, depressurize and drain the system.
- Flush the assembly thoroughly before starting work, as it may contain medium residues.

CAUTION

Risk to health due to contact with the process medium!

- Pay attention to the warnings in the safety data sheets.
- Wear protective gloves, protective goggles and protective clothing, particularly when working with reagents, chemicals or process solutions.
- In case of contact with eyes or skin, rinse the affected area with plenty of water and then seek medical advice. Show the relevant safety data sheet to the physician.

NOTICE**Material damage due to solvents!**

- Do not use any halogen-containing organic solvents or acetone. These solvents could destroy plastic parts of the device.
- The Operating Instructions for the device must be observed before using solvents.

 Potential impact on the process

Before decommissioning an active device, the potential impact on the overall process must be taken into account! This applies in particular when using the switching contacts, the analog signal outputs or the communication interface of the associated measuring instrument to control process variables. Coordinate service tasks with the operator!

5 Additional documentation

Detailed information on the device can be found in the Operating Instructions for the assembly and in the other documentation available at:

- CYA10 product page: <https://www.endress.com/cya10>
- Spare parts finder: <https://sft-p.cdn.endress.com/>
- www.endress.com/device-viewer
- Smartphone/tablet: Endress+Hauser Operations app

6 Disposal

 If required by the Directive 2012/19/EU on waste electrical and electronic equipment (WEEE), the product is marked with the depicted symbol in order to minimize the disposal of WEEE as unsorted municipal waste. Do not dispose of products bearing this marking as unsorted municipal waste. Instead, return them to the manufacturer for disposal under the applicable conditions.



71754783

www.addresses.endress.com
