

FDA- Konformitätserklärung

Produkt

Product

Wir als Hersteller von Produkten für die Analyse erklären, dass die prozessmedienberührenden Komponenten des oben genannten Produktes in ihrer Zusammensetzung der unten aufgeführten Bestimmung der

„Food and Drug Administration (FDA)“

entsprechen.

Diese Herstellererklärung ist nur gültig für das oben genannte Produkt.

| Komponente Component | Material Material | FDA-Konformität FDA conformity |
|---------------------------------|---------------------------------------|-----------------------------------|
| Formdichtungen Moulded seals | EPDM | 21 CFR 177.2600 |
| O-Ringe O-rings | E70Q | |
| | EPDM | 21 CFR 177.2600 |
| Fett Grease | E70Q | |
| | Grease based on silicone oil and PTFE | 21 CFR 178.3570 |
| | PARALIQ GTE 703 | |

Endress+Hauser Conducta GmbH+Co. KG
Dieselstraße 24, 70839 Gerlingen, Germany

Datum

Date

21.01.2019



i.V. Uwe Rößiger
Manager Certifications & Approvals



i.V. Janina Haschke
Certifications & Approvals

Zertifikats-Nr. | Certificate-no.
FDA_00204_01.19

FDA Declaration of Conformity

XPC0001-A QA

Kit, seals EPDM FDA, G1 1/4 wetted parts, single chamber

We as product manufacturer of analysis products declare that the following type of product consists wetted parts with process contact, which in their composition are in compliance with below mentioned regulation

Food and Drug Administration (FDA).

This manufacturer's declaration is only valid for above mentioned product.