

FDA- Konformitätserklärung

Produkt:

**Dichtungssatz Silikon CPA442, FDA
51507171**

Wir als Hersteller von Produkten für die Analyse erklären, dass die prozessmedienberührenden Komponenten des oben genannten Produktes in ihrer Zusammensetzung der unten aufgeführten Bestimmung der

„Food and Drug Administration (FDA)“

entsprechen.

Diese Herstellererklärung ist nur gültig für das oben genannte Produkt.

Komponente

Component

Formdichtung & O-Ring
Moulded seal & O-ring

Material

Material

VMQ Silikon
VMQ Silicon

FDA-Konformität

FDA conformity

21 CFR 177.2600 (a – d)

Endress+Hauser Conducta GmbH+Co. KG
Dieselstraße 24, 70839 Gerlingen, Germany


Datum

Date

21.10.2019



i.V. Uwe Rößiger
Manager Certifications & Approvals



i.A. Julia Groß
Specialist R&D Chemical Sensors

Zertifikats-Nr. | Certificate-no.
FDA_00227_01.19

Endress + Hauser 

People for Process Automation

FDA Declaration of Conformity

Product:

**Sealing set Silicon CPA442, FDA listed
51507171**

We as product manufacturer of analysis products declare that the following type of product consists wetted parts with process contact, which in their composition are in compliance with below mentioned regulation

Food and Drug Administration (FDA).

This manufacturer's declaration is only valid for above mentioned product.