

Installation Instructions

Replacing cover, O-ring cover, Interface gasket

Promag, Promass, Prosonic Flow 300, 500,
t-mass 300



Replacement of cover, O-ring, interface seal

Promag, Promass, Prosonic Flow 300, 500, t-mass 300

Table of contents

1	Overview of spare part sets	4
2	Designated use	6
3	Personnel authorized to carry out repairs	6
4	Safety instructions	7
5	Symbols used	8
6	Tools list	9
7	Connection compartment cover non-Ex/Exde and connection compartment cover Exd	9
8	Promag, Promass, Prosonic Flow, t-mass 300	10
9	Promag, Promass, Prosonic Flow 300, 500, t-mass 300	11
10	Promag, Promass 300	16
11	Disposal	24

1 Overview of spare part sets

The Installation Instructions apply to the following spare part sets:

Order number	Original spare part set	Content
71323016	Cover set, window, aluminum, non-Ex/Exde	1 × complete cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323017	Cover set, glass, aluminum, -50 °C, non-Ex/Exde	1 × complete cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323018	Cover set, blind, aluminum, non-Ex/Exde	1 × cover , 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323019	Cover set, blind, aluminum, -50 °C, non-Ex/Exde	1 × cover , 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323020	Cover set, window, aluminum, Exd	1 × complete cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323021	Cover set, window, aluminum, -50 °C, Exd	1 × complete cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323022	Cover set, blind, aluminum, Exd	1 × cover , 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323023	Cover set, blind, aluminum, -50 °C, Exd	1 × cover , 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323024	Cover set, window, cast, Ex	1 × complete cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323025	Cover set, window, cast, -50 °C, Ex	1 × complete cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323026	Cover seal set, window/blind	10 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323027	Cover seal set, window/blind, -50 °C	1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323028	Cover set, electronics compartment, aluminum, non-Ex	1 × cover, 1 × O-ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323029	Cover set, electronics compartment, aluminum, -50 °C, Non-Ex	1 × cover, 1 × O-ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323030	Cover seal set, electronics compartment, non-Ex	10 × O-ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323031	Cover seal set, electronics compartment, -50 °C, Non-Ex	1 × O-ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323032	Cover set, electronics compartment, aluminum, Ex	1 × cover, 1 × O-ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323033	Cover set, electronics compartment, aluminum, -50 °C, Ex	1 × cover, 1 × O-ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323034	Cover set, electronics compartment, cast, Ex	1 × cover, 1 × O-ring (120.2 × 3.53 mm/4.7 × 1.4 in)

Order number	Original spare part set	Content
71323035	Cover set, electronics compartment, cast, -50 °C, Ex	1 × cover, 1 × O-ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323036	Cover seal set, electronics compartment, Ex	10 × cover, 1 × O-ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323037	Cover seal set, electronics compartment, -50 °C, Ex	1 × cover, 1 × O-ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323069	Interface seal set, Promag 300, 500	5 × O-ring (88.50 × 3.53/4.7 × 1.4 in), 5 × O-ring (29.87 × 1.78/1.2 × 0.1 in)
71323070	Interface seal set, Promag 500, -50 °C	1 × O-ring (88.50 × 3.53/4.7 × 1.4 in), 1 × O-ring (29.87 × 1.78/1.2 × 0.1 in)
71323071	Interface seal set, 300, 500	5 × O-ring (88.50 × 3.53/4.7 × 1.4 in)
71323072	Interface seal set, 300, 500, -50 °C	1 × O-ring (88.50 × 3.53/4.7 × 1.4 in)
71327102	Cover set, blind, cast, Ex	1 × complete cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in) 1 × blind front membrane (Ø 79.00 mm/3.1 in)
71327103	Cover set, blind, cast, -50 °C, Ex	1 × cover, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in) 1 × blind front membrane (Ø 79.00 mm/3.1 in)
71368481	Cover set, glass, hygienic, Z2	1 × cover with glass, 1 × seal to connection compartment
71368482	Cover set, polycarbonate, hygienic, non-Ex/ Div.2	1 × cover, 1 × seal to connection compartment
71368483	Cover set, blind, hygienic	1 × cover, 1 × seal to connection compartment
71368484	Cover seal set, glass/polycarbonate/blind, hygienic	10 × seal to connection compartment
71368486	Cover set, electronics compartment, hygienic	1 × cover, 1 × seal to electronics compartment
71368487	Cover seal set, electronics compartment, hygienic	10 × seal to electronics compartment
71368490	Interface seal set, Promag 300, hygienic	5 × O-ring 88.50 × 3.53 mm/3.5 × 0.14 in), 5 × seal, 5 × O-ring 29.87 × 1.78 mm/1.2 × 0.1 in)
71368491	Interface seal set, Promass 300, hygienic	5 × O-ring 88.50 × 3.53 mm/3.5 × 0.14 in), 5 × seal
71368492	Interface seal set, Promass 300, hygienic, -50°C	1 × O-ring 88.50 × 3.53 mm/3.5 × 0.14 in), 1 × seal




- The order number of the spare part set (on the product label on the package) can differ from the production number (on the label directly on the spare part)!
- You can find the order number of the relevant spare part set by entering the production number of the spare part in the spare part search tool.
- We recommend that you keep the Installation Instructions and packaging together at all times.

2 Designated use

The spare part sets and Installation Instructions are used to replace a faulty unit with a functioning unit of the same type. Only original parts from Endress+Hauser may be used. Only spare parts kits designed by Endress+Hauser for the measuring device can be used at any time.

Inspection is performed using the W@M Device Viewer. The procedure involved is described as follows.

i A spare parts overview is located in the interior of some measuring devices. If the spare parts kit is specified on this list, no inspection is required.



3 Personnel authorized to carry out repairs

Authorization to carry out repairs depends on the measuring device's approval type. The table below shows the authorized group of people in each case.

i Whoever carries out the repairs has full responsibility to ensure that work is carried out safely and to the required quality standard. He/she must also guarantee the safety of the device following repair.

Measuring device approval	Spare part	Personnel authorized to perform repairs ¹⁾
Without approval With approval (e.g. IECEx)	Cover, cover seal	1, 2, 3
Without approval With approval (e.g. IECEx)	Interface seal	2, 3
For custody transfer	Cover, cover seal, interface seal	4

1) 1 = Qualified specialist on customer side, 2 = Service technician authorized by Endress+Hauser,

3 = Endress+Hauser (return measuring device to manufacturer)

4 = Check with local approval center if installation/alteration must be performed under supervision.

4 Safety instructions

- Check whether the spare part matches the labeling on the measuring device as described on the cover page.
- The spare part sets and Installation Instructions are used to replace a faulty unit with a functioning unit of the same type.
Only use original parts from Endress+Hauser.
- Comply with national regulations governing mounting, electrical installation, commissioning, maintenance and repair procedures.
- The following requirements must be met with regard to specialized technical staff for the mounting, electrical installation, commissioning, maintenance and repair of the measuring devices:
 - Specialized technical staff must be trained in instrument safety.
 - They must be familiar with the individual operating conditions of the devices.
 - In the case of Ex-certified measuring devices, they must also be trained in explosion protection.
- The measuring device is energized! Risk of fatal injury from electric shock. Open the measuring device only when the device is de-energized.
- For measuring devices intended for use in hazardous locations, please observe the guidelines in the Ex documentation (XA).
- In the case of measuring devices in safety-related applications in accordance with IEC 61508 or IEC 61511: Commission in accordance with Operating Instructions after repair. Document the repair procedure.
- Before removing the device: set the process to a safe state and purge the pipe of dangerous process substances.
- Hot surfaces! Risk of injury! Before commencing work: allow the system and measuring device to cool down to a touchable temperature.
- In the case of devices in custody transfer, the custody transfer status no longer applies once the seal has been removed.
- The Operating Instructions for the device must be followed.
- Risk of damaging the electronic components! Ensure you have a working environment protected from electrostatic discharge.
- After removing the electronics compartment cover: risk of electrical shock due to missing touch protection!
Turn the measuring device off before removing internal covers.
- Modifications to the measuring device are not permitted.
- Only open the housing for a brief period. Avoid the penetration of foreign bodies, moisture or contaminants.
- Replace defective seals only with original seals from Endress+Hauser.
- If threads are damaged or defective, the measuring device must be repaired.
- Threads (e.g. of the electronics compartment cover and connection compartment cover) must be lubricated if an abrasion-proof dry lubricant is not available. Use acid-free, non-hardening lubricant.

- If spacing is reduced or the dielectric strength of the measuring device cannot be guaranteed during repair work, perform a test on completion of the work (e.g. high-voltage test in accordance with the manufacturer's instructions).
- Service plug:
 - Do not connect in explosive atmospheres.
 - Only connect to Endress+Hauser service devices.
- Observe the instructions for transporting and returning the device outlined in the Operating Instructions.

 If you have any questions, please contact your [Endress+Hauser service organization](#).

5 Symbols used

5.1 Safety symbols

DANGER

This symbol alerts you to a dangerous situation. Failure to avoid this situation will result in serious or fatal injury.

WARNING

This symbol alerts you to a dangerous situation. Failure to avoid this situation can result in serious or fatal injury.

CAUTION

This symbol alerts you to a dangerous situation. Failure to avoid this situation can result in minor or medium injury.

NOTICE

This symbol contains information on procedures and other facts which do not result in personal injury.

5.2 Symbols for certain types of information

Symbol	Meaning
	Permitted Procedures, processes or actions that are permitted.
	Forbidden Procedures, processes or actions that are forbidden.
	Tip Indicates additional information.
1, 2, 3...	Series of steps

6 Tools list




7 Connection compartment cover non-Ex/Exde and connection compartment cover Exd

⚠ CAUTION

For transmitter aluminum housing/stainless steel cast housing: the threads of the non-Ex and /Exde housing and the Exd housing are a different length.

- Make sure that the covers are not accidentally interchanged!




1 Engineering units mm (in)

- A Connection compartment cover non-Ex/Exde
 B Connection compartment cover Exd

8 Promag, Promass, Prosonic Flow, t-mass 300

8.1 Aluminum transmitter housing, non-Ex, Ex; stainless steel cast housing, Ex

8.1.1 Replacing the blind connection compartment cover, replacing the electronics compartment cover, replacing the O-ring for the cover




8.1.2 Reassembling the seal and cover

Reassembly is carried out in reverse order.

9 Promag, Promass, Prosonic Flow 300, 500, t-mass 300

9.1 Aluminum transmitter housing, non-Ex, Ex; stainless steel cast housing, Ex


9.1.1 Replacing the connection compartment cover with window, replacing the electronics compartment cover, replacing the O-ring for the cover













9.1.2 Reassembling the seal and cover

Reassembly is carried out in reverse order.


9.2 Replacing the interface seal and the interface O-ring









- ▶ Removal of transmitter version for non-hazardous areas: disconnect the ribbon cable, loosen the screws and turn the transmitter and lift it up.



- ▶ Removal of transmitter version for hazardous areas: loosen the screws and turn the transmitter and lift it up.






- ▶ Promass, Prosonic Flow, 300,
500, t-mass 300

9.2.1 Reassembling the transmitter housing

Reassembly is carried out in reverse order. Note the following:


9.2.2 Note on reassembling the Promag 300

- i** Set the grounding switch of the ISEM electronics module to the position on the "left" (closed), as illustrated in the diagram below.



9.2.3 Note on reassembling the Promag 500


i Set the grounding switch of the ISEM electronics module to the position on the "right" (open), as illustrated in the diagram below.



9.2.4 Note on reassembling the Promag, Promass, Prosonic Flow 300, 500, t-mass 300


i Check if the S-DAT is inserted in the amplifier. Otherwise, check if the S-DAT is inserted in the sensor interface.

9.2.5 Installing the transmitter housing for non-hazardous areas



- ▶ Installation of the transmitter version for non-hazardous areas: fit the transmitter in place, turn it and tighten the screws.

9.2.6 Installing the transmitter housing for hazardous areas



- ▶ Installation of the transmitter version for hazardous areas: fit the transmitter in place, turn it and tighten the screws.


For detailed information on the electrical connection, see the "Electrical connection" section of the Operating Instructions for the device.

10 Promag, Promass 300


10.1 Stainless steel transmitter housing, hygienic non-Ex

10.1.1 Releasing the cover lock for transmitter in Zone 2

A cover lock is located on the connection compartment cover and the electronics compartment cover on devices for Zone 2. The cover lock must be released before the connection compartment cover/electronics compartment cover can be unscrewed.



10.1.2 Replacing the blind connection compartment cover, replacing the seal for the cover




10.1.3 Reassembling the connection compartment cover and seal


Reassembly is carried out in reverse order. Note the following:



Make sure that the seal is mounted in the right position, see the graphic below.




10.1.4 Replacing the connection compartment cover with window, replacing the seal for the cover




10.1.5 Reassembling the connection compartment cover and cover

Reassembly is carried out in reverse order. Note the following:

- Make sure that the seal is mounted in the right position, see the graphic below.






10.1.6 Replacing the electronics compartment cover, replacing the seal for the cover






10.1.7 Reassembling the electronics compartment cover and seal









Reassembly is carried out in reverse order.

10.1.8 Replacing the interface seal and the interface O-ring




► Remove the display module if present.





▶ Promag 300




► Promass 300

10.1.9 Reassembling the transmitter housing

Reassembly is carried out in reverse order. Note the following:



10.1.10 Note on reassembling the Promag 300

- i** Set the grounding switch of the ISEM electronics module to the position on the "left" (closed), as illustrated in the diagram below.





10.1.11 Note on reassembling the Promag, Promass 300


i Check if the S-DAT is inserted in the amplifier (Fig. 1 below). Otherwise, check if the S-DAT is inserted in the sensor interface (Fig. 2 below).




i Otherwise, remove the S-DAT from the defective amplifier (Fig. 1 below) and install it in the new amplifier (Fig. 2 or 3 below).




► **For Promag:** Insert S-DAT into the ISEM electronics module.



► **For Promass:** Insert S-DAT into the ISEM electronics module.




- ▶ Align and position the ISEM electronics module in the transmitter housing in accordance with the diagram.



- ▶ Press down the ISEM electronics module (1), tighten screws (2) and insert cable (3).

10.1.12 Installing the transmitter housing for non-hazardous areas



For detailed information on the electrical connection, see the "Electrical connection" section of the Operating Instructions for the device.

11 Disposal



If required by the Directive 2012/19/EU on waste electrical and electronic equipment (WEEE), the product is marked with the depicted symbol in order to minimize the disposal of WEEE as unsorted municipal waste. Do not dispose of products bearing this marking as unsorted municipal waste. Instead, return them to Endress+Hauser for disposal under the applicable conditions.

Austausch Deckel, O-Ring, Schnittstellendichtung

Promag, Promass, Prosonic Flow 300, 500, t-mass 300

Inhaltsverzeichnis

1	Übersicht Ersatzteilsets	26
2	Bestimmungsgemäße Verwendung	28
3	Reparaturberechtigte Personen	28
4	Sicherheitshinweise	29
5	Verwendete Symbole	30
6	Werkzeugliste	31
7	Anschlussraumdeckel Ex-frei/Exde und Anschlussraumdeckel Exd	31
8	Promag, Promass, Prosonic Flow, t-mass 300	32
9	Promag, Promass, Prosonic Flow 300, 500, t-mass 300	32
10	Promag, Promass 300	38
11	Entsorgung	46

1 Übersicht Ersatzteilsets

Die Einbauanleitung ist für folgende Ersatzteilsets gültig:

Bestellnummer	Original Ersatzteilset	Inhalt
71323016	Set Deckel, Schauglas, Alu, Ex-frei/Exde	1 × Deckel komplett, 1 × O-Ring (110,70 × 3,53 mm/4,4 × 1,4 in)
71323017	Set Deckel, Glas, Alu, -50 °C, Ex-frei/Exde	1 × Deckel komplett, 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323018	Set Deckel, blind, Alu, Ex-frei/Exde	1 × Deckel , 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323019	Set Deckel, blind, Alu,-50 °C, Ex-frei/Exde	1 × Deckel , 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323020	Set Deckel, Schauglas, Alu, Exd	1 × Deckel komplett, 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323021	Set Deckel, Schauglas, Alu, -50 °C, Exd	1 × Deckel komplett, 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323022	Set Deckel, blind, Alu, Exd	1 × Deckel , 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323023	Set Deckel, blind, Alu, -50 °C, Exd	1 × Deckel , 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323024	Set Deckel, Schauglas, Guss, Ex	1 × Deckel komplett, 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323025	Set Deckel, Schauglas, Guss, -50 °C, Ex	1 × Deckel komplett, 1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323026	Set Deckeldichtung, Schauglas/Blind	10 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323027	Set Deckeldichtung, Schauglas/Blind, -50 °C	1 × O-Ring (110.70 × 3.53 mm/4.4 × 1.4 in)
71323028	Set Deckel, Elektronikraum, Alu, Ex-frei	1 × Deckel, 1 × O-ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323029	Set Deckel, Elektronikraum, Alu, -50 °C, Ex-frei	1 × Deckel, 1 × O-ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323030	Set Deckeldichtung, Elektronikraum, Ex-frei	10 × O-Ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323031	Set Deckeldichtung, Elektronikraum, -50 °C, Ex-frei	1 × O-Ring (117.10 × 3.53 mm/4.6 × 1.4 in)
71323032	Set Deckel, Elektronikraum, Alu, Ex	1 × Deckel, 1 × O-Ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323033	Set Deckel, Elektronikraum, Alu,-50 °C, Ex	1 × Deckel, 1 × O-Ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323034	Set Deckel, Elektronikraum, Guss, Ex	1 × Deckel, 1 × O-Ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323035	Set Deckel, Elektronikraum,Guss,-50 °C, Ex	1 × Deckel, 1 × O-Ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323036	Set Deckeldichtung, Elektronikraum, Ex	10 × Deckel, 1 × O-Ring (120.2 × 3.53 mm/4.7 × 1.4 in)

Bestellnummer	Original Ersatzteilset	Inhalt
71323037	Set Deckeldichtung, Elektronikraum, -50 °C, Ex	1 × Deckel, 1 × O-Ring (120.2 × 3.53 mm/4.7 × 1.4 in)
71323069	Set Dichtung Schnittstelle, Promag 300, 500	5 × O-Ring (88.50 × 3.53/4.7 × 1.4 in), 5 × O-Ring (29.87 × 1.78/1.2 × 0.1 in)
71323070	Set Dichtung Schnittstelle, Promag 500, -50 °C	1 × O-Ring (88.50 × 3.53/4.7 × 1.4 in), 1 × O-Ring (29.87 × 1.78/1.2 × 0.1 in)
71323071	Set Dichtung Schnittstelle, 300, 500	5 × O-Ring (88.50 × 3.53/4.7 × 1.4 in)
71323072	Set Dichtung Schnittstelle, 300, 500, -50 °C	1 × O-Ring (88.50 × 3.53/4.7 × 1.4 in)
71327102	Set Deckel, blind, Guss, Ex	1 × Deckel komplett, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in) 1 × Frontfolie blind (Ø 79,00 mm/3.1 in)
71327103	Set Deckel, blind, Guss, -50 °C, Ex	1 × Deckel, 1 × O-ring (110.70 × 3.53 mm/4.4 × 1.4 in) 1 × Frontfolie blind (Ø 79,00 mm/3.1 in)
71368481	Set Deckel, Glas, hygienisch, Z2	1 × Deckel mit Glas, 1 × Dichtung zu Anschlussraum
71368482	Set Deckel, Polycarbonat, hygienisch, Ex-frei/Div.2	1 × Deckel, 1 × Dichtung zu Anschlussraum
71368483	Set Deckel, blind, hygienisch	1 × Deckel, 1 × Dichtung zu Anschlussraum
71368484	Set Deckeldichtung, Glas/Polycarbonat/blind, hygienisch	10 × Dichtung zu Anschlussraum
71368486	Set Deckel, Elektronikraum, hygienisch	1 × Deckel, 1 × Dichtung zu Elektronikraum
71368487	Set Deckeldichtung, Elektronikraum, hygienisch	10 × Dichtung zu Elektronikraum
71368490	Set Dichtung Schnittstelle, Promag 300, hygienisch	5 × O-Ring 88,50 × 3,53 mm/3,5 × 0,14 in), 5 × Dichtung, 5 × O-Ring 29,87 × 1,78 mm/1,2 × 0,1 in)
71368491	Set Dichtung Schnittstelle, Promass 300, hygienisch	5 × O-Ring 88,50 × 3,53 mm/3,5 × 0,14 in), 5 × Dichtung
71368492	Set Dichtung Schnittstelle, Promass 300, hygienisch, -50°C	1 × O-Ring 88,50 × 3,53 mm/3,5 × 0,14 in), 1 × Dichtung




- Die Bestellnummer des Ersatzteilsets (auf dem Produktaufkleber der Verpackung) kann sich von der Produktionsnummer (auf dem Aufkleber direkt auf dem Ersatzteil) unterscheiden!
- Durch Eingabe der Produktionsnummer des Ersatzteiles im Ersatzteifindetool kann die Bestellnummer des entsprechenden Ersatzteilsets ermittelt werden.
- Wir empfehlen Einbauanleitung und Verpackung immer zusammen aufzubewahren.

2 Bestimmungsgemäße Verwendung

Ersatzteilset und Einbauanleitung dienen dazu, eine defekte Einheit gegen eine funktionierende Einheit des gleichen Typs zu ersetzen. Es dürfen nur Originalteile von Endress+Hauser verwendet werden. Grundsätzlich dürfen nur Ersatzteilsets verwendet werden, die von Endress+Hauser für das Messgerät vorgesehen sind.

Die Überprüfung ist via W@M Device Viewer durchzuführen, die Vorgehensweise dazu ist nachfolgend beschrieben.

- i** Bei einigen Messgeräten befindet sich im Inneren des Gerätes eine Ersatzteilübersicht. Ist das Ersatzteilset dort aufgelistet, entfällt die Überprüfung.



3 Reparaturberechtigte Personen

Die Berechtigung zur Durchführung einer Reparatur ist von der Zulassung des Messgeräts abhängig. Die Tabelle zeigt den jeweils berechtigten Personenkreis.

- i** Die Person, die eine Reparatur vornimmt, übernimmt die Verantwortung für die Sicherheit während der Arbeiten, die Qualität der Ausführung und die Sicherheit des Geräts nach der Reparatur.

Zulassung des Messgeräts	Ersatzteil	Reparaturberechtigter Personenkreis ¹⁾
ohne Zulassung mit Zulassung (z.B. IECEEx)	Deckel, Deckeldichtung	1, 2, 3
ohne Zulassung mit Zulassung (z.B. IECEEx)	Dichtung Schnittstelle	2, 3
Bei eichfähigem Verkehr	Deckel, Deckeldichtung, Dichtung Schnittstelle	4

1) 1 = Ausgebildete Fachkraft des Kunden, 2 = Von Endress+Hauser autorisierter Servicetechniker,
3 = Endress+Hauser (Messgerät an Hersteller zurücksenden)
4 = Mit der lokalen Zulassungsstelle prüfen, ob ein Ein-/Umbau unter Aufsicht erfolgen muss.

4 Sicherheitshinweise

- Prüfen, ob das vorliegende Ersatzteil zur Kennzeichnung auf dem Messgerät passt, wie auf der Titelseite beschrieben.
- Ersatzteilset und Einbauanleitung dienen dazu, eine defekte Einheit gegen eine funktionierende Einheit des gleichen Typs zu ersetzen.
Nur Originalteile von Endress+Hauser verwenden.
- Nationale Vorschriften bezüglich der Montage, elektrischen Installation, Inbetriebnahme, Wartung und Reparatur einhalten.
- Folgende Anforderungen an das Fachpersonal für Montage, elektrische Installation, Inbetriebnahme, Wartung und Reparatur der Messgeräte müssen erfüllt sein:
 - In Gerätesicherheit ausgebildet.
 - Mit den jeweiligen Einsatzbedingungen der Geräte vertraut.
 - Bei Ex-zertifizierten Messgeräten: zusätzlich im Explosionsschutz ausgebildet.
- Messgerät unter Spannung! Lebensgefahr durch Stromschlag. Messgerät nur im spannungslosen Zustand öffnen.
- Bei Messgeräten für den explosionsgefährdeten Bereich: Hinweise in der Ex-Dokumentation (XA) beachten.
- Bei Messgeräten in sicherheitstechnischen Applikationen gemäß IEC 61508 bzw. IEC 61511: Nach Reparatur Neuinbetriebnahme gemäß Betriebsanleitung durchführen. Reparatur dokumentieren.
- Vor einem Geräteausbau: Prozess in sicheren Zustand bringen und Leitung von gefährlichen Prozessstoffen befreien.
- Verbrennungsgefahr durch heiße Oberflächen! Vor Arbeitsbeginn: Anlage und Messgerät auf berührungssichere Temperatur abkühlen.
- Bei Messgeräten im abrechnungspflichtigen Verkehr: Nach Entfernen der Plombe ist der geeichte Zustand aufgehoben.
- Die Betriebsanleitung zum Messgerät ist zu beachten.
- Beschädigungsgefahr elektronischer Bauteile! Eine ESD-geschützte Arbeitsumgebung herstellen.
- Nach Entfernen der Elektronikabdeckung: Stromschlaggefahr durch aufgehobenen Berührungsschutz!
Messgerät ausschalten, bevor interne Abdeckungen entfernt werden.
- Änderungen am Messgerät sind nicht zulässig.
- Gehäuse nur kurzzeitig öffnen. Eindringen von Fremdkörpern, Feuchtigkeit oder Verunreinigung vermeiden.
- Defekte Dichtungen nur durch Original-Dichtungen von Endress+Hauser ersetzen.
- Defekte Gewinde erfordern eine Instandsetzung des Messgeräts.
- Gewinde (z.B. von Elektronikraum- und Anschlussraumdeckel) müssen geschmiert sein, sofern keine abriebfeste Trockenschmierung vorhanden ist. Säurefreies, nicht härzendes Fett verwenden.

- Wenn bei den Reparaturarbeiten Abstände reduziert oder die Spannungsfestigkeit des Messgeräts nicht sichergestellt werden kann: Prüfung nach Abschluss der Arbeiten durchführen (z.B. Hochspannungstest gemäß Herstellerangaben).
- Servicestecker:
 - Nicht in explosionsfähiger Atmosphäre anschließen.
 - Nur an Servicegeräte von Endress+Hauser anschließen.
- Die in der Betriebsanleitung aufgeführten Hinweise zum Transport und zur Rücksendung beachten.

 Bei Fragen kontaktieren Sie bitte Ihre zuständige [Endress+Hauser Serviceorganisation](#).

5 Verwendete Symbole

5.1 Warnhinweissymbole

GEFAHR

Dieser Hinweis macht auf eine gefährliche Situation aufmerksam, die, wenn sie nicht vermieden wird, zu Tod oder schwerer Körperverletzung führen wird.

WARNUNG

Dieser Hinweis macht auf eine gefährliche Situation aufmerksam, die, wenn sie nicht vermieden wird, zu Tod oder schwerer Körperverletzung führen kann.

VORSICHT

Dieser Hinweis macht auf eine gefährliche Situation aufmerksam, die, wenn sie nicht vermieden wird, zu leichter oder mittelschwerer Körperverletzung führen kann.

HINWEIS

Dieser Hinweis enthält Informationen zu Vorgehensweisen und weiterführenden Sachverhalten, die keine Körperverletzung nach sich ziehen.

5.2 Symbole für Informationstypen

Symbol	Bedeutung
	Erlaubt Abläufe, Prozesse oder Handlungen, die erlaubt sind.
	Verboten Abläufe, Prozesse oder Handlungen, die verboten sind.
	Tipp Kennzeichnet zusätzliche Informationen.
	Handlungsschritte

6 Werkzeugliste




7 Anschlussraumdeckel Ex-frei/Exde und Anschlussraumdeckel Exd

⚠ VORSICHT

Für Messumformer Aluminiumgehäuse/Gussgehäuse rostfrei: Die Gewindelänge zwischen dem Ex-freien- und /Exde- Gehäuse und dem Exd-Gehäuse unterscheiden sich in der Länge.

► Darauf achten, dass die Deckel nicht vertauscht werden!




2 Masseinheiten mm (in)

- A Anschlussraumdeckel Ex-frei/Exde
 B Anschlussraumdeckel Exd

8 Promag, Promass, Prosonic Flow, t-mass 300

8.1 Messumformergehäuse Aluminium Ex-frei, Ex, Guss rostfrei Ex

8.1.1 Austausch Anschlussraumdeckel blind, Austausch Elektronikraumdeckel, Austausch O-Ring zu Deckel




8.1.2 Zusammenbau Dichtung und Deckel

Der Zusammenbau erfolgt in umgekehrter Reihenfolge.

9 Promag, Promass, Prosonic Flow 300, 500, t-mass 300

9.1 Messumformergehäuse Aluminium Ex-frei, Ex, Guss rostfrei Ex


9.1.1 Austausch Anschlussraumdeckel mit Sichtglas, Austausch Elektronikraumdeckel, Austausch O-Ring zu Deckel




9.1.2 Zusammenbau Dichtung und Deckel


Der Zusammenbau erfolgt in umgekehrter Reihenfolge.

9.2 Austausch Dichtung Schnittstelle und O-Ring Schnittstelle






► Ausbau Messumformer in nicht explosionsgeschützter Ausführung:
Flachbandkabel abziehen, Schrauben lösen und Messumformer drehen
und hochziehen.



► Ausbau Messumformer in explosionsgeschützter Ausführung: Schrauben lösen und Messumformer drehen und hochziehen.

► Promag 300, 500




► Promass, Prosonic Flow, 300,
500, t-mass 300

9.2.1 Zusammenbau Messumformergehäuse

Der Zusammenbau erfolgt in umgekehrter Reihenfolge. Folgendes ist zu beachten:


9.2.2 Hinweis zum Zusammenbau Promag 300

i Erdungsschalter des ISEM-Elektronikmodul auf Position "links" stellen (geschlossen), wie im Bild unten gezeigt.



9.2.3 Hinweis zum Zusammenbau Promag 500


i Erdungsschalter des ISEM-Elektronikmodul auf Position "rechts" stellen (offen), wie im Bild unten gezeigt.



9.2.4 Hinweis zum Zusammenbau Promag, Promass, Prosonic Flow 300, 500, t-mass 300


i Kontrollieren, ob der S-DAT im Messverstärker eingesteckt ist. Andernfalls kontrollieren, ob der S-DAT in der Sensorschnittstelle eingesteckt ist.

9.2.5 Einbau Messumformergehäuse in nicht explosionsgeschützter Ausführung



- ▶ Einbau Messumformer in nicht explosionsgeschützter Ausführung:
Messumformer aufsetzen und drehen, Schrauben anziehen.

9.2.6 Einbau Messumformergehäuse in explosionsgeschützter Ausführung



- ▶ Einbau Messumformer in explosionsgeschützter Ausführung:
Messumformer aufsetzen und drehen, Schrauben anziehen.


Detaillierte Angaben zum elektrischen Anschluss: Kapitel "Elektrischer Anschluss", Betriebsanleitung zum Gerät.

10 Promag, Promass 300


10.1 Messumformergehäuse rostfrei hygienisch Ex-frei

10.1.1 Lösen der Deckelsicherung für Messumformer in Zone 2

 Bei Geräten für Zone 2 befindet sich am Anschlussraumdeckel und am Elektronikraumdeckel eine Deckelsicherung. Die Deckelsicherung muss gelöst werden, bevor man den Anschlussraumdeckel/Elektronikraumdeckel abschrauben kann.




10.1.2 Austausch Anschlussraumdeckel blind, Austausch Dichtung zu Deckel




10.1.3 Zusammenbau Anschlussraumdeckel und Dichtung

Der Zusammenbau erfolgt in umgekehrter Reihenfolge. Folgendes ist zu beachten:

- i** Darauf achten, dass die Dichtung in der richtigen Position montiert wird, siehe im Bild unten.




10.1.4 Austausch Anschlussraumdeckel mit Sichtglas, Austausch Dichtung zu Deckel




10.1.5 Zusammenbau Anschlussraumdeckel und Deckel

Der Zusammenbau erfolgt in umgekehrter Reihenfolge. Folgendes ist zu beachten:

-  Darauf achten, dass die Dichtung in der richtigen Position montiert wird, siehe im Bild unten.






10.1.6 Austausch Elektronikraumdeckel, Austausch Dichtung zu Deckel






10.1.7 Zusammenbau Elektronikraumdeckel und Dichtung








Der Zusammenbau erfolgt in umgekehrter Reihenfolge.

10.1.8 Austausch Dichtung Schnittstelle und O-Ring Schnittstelle




► Anzeigemodul entfernen, falls vorhanden.





▶ Promag 300




► Promass 300

10.1.9 Zusammenbau Messumformergehäuse

Der Zusammenbau erfolgt in umgekehrter Reihenfolge. Folgendes ist zu beachten:


10.1.10 Hinweis zum Zusammenbau Promag 300

 Erdungsschalter des ISEM-Elektronikmodul auf Position "links" stellen (geschlossen), wie im Bild unten gezeigt.




10.1.11 Hinweis zum Zusammenbau Promag, Promass 300

- i** Kontrollieren, ob der S-DAT im Messverstärker eingesteckt ist (Bild 1 unten).
Andernfalls kontrollieren, ob der S-DAT in der Sensorschnittstelle eingesteckt ist (Bild 2 unten).




- i** Ansonsten den S-DAT aus dem defekten Messverstärker demontieren (Bild 1 unten) und in den neuen Messverstärker montieren (Bild 2 oder 3 unten).




► Für Promag: S-DAT in das ISEM-Elektronikmodul einstecken.

► Für Promass: S-DAT in das ISEM-Elektronikmodul einstecken.




- ISEM-Elektronikmodul gemäss Abbildung im Messumformergehäuse ausrichten und platzieren.



- ISEM-Elektronikmodul herunter drücken (1), Schrauben fest ziehen (2) und Kabel einstecken (3).

10.1.12 Einbau Messumformergehäuse in nicht explosionsgeschützter Ausführung



Detaillierte Angaben zum elektrischen Anschluss: Kapitel "Elektrischer Anschluss", Betriebsanleitung zum Gerät.

11 Entsorgung



Gemäß der Richtlinie 2012/19/EU über Elektro- und Elektronik-Altgeräte (WEEE) ist das Produkt mit dem abgebildeten Symbol gekennzeichnet, um die Entsorgung von WEEE als unsortierten Hausmüll zu minimieren. Gekennzeichnete Produkte nicht als unsortierter Hausmüll entsorgen, sondern zu den gültigen Bedingungen an Endress+Hauser zurückgeben.



71521752

www.addresses.endress.com
