



Certificate of Compliance

Certificate: 70134179 **Master Contract:** 205557 (205557)

Project: 70185140 **Date Issued:** 2018-07-16

Issued to: Endress+Hauser Conducta GmbH & Co. KG
Dieselstrasse 24
Gerlingen, 70839
GERMANY
Attention: Uwe Roessiger

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.



Issued by: *Dimcho Genov*
Dimcho Genov

PRODUCTS

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT-Certified to US Standards

Automated Sensor Cleaning and Calibration System, Liquiline Control CDC90

Rated 100 – 230 Vac, 50/60Hz, 50VA max;

One set of relay contacts rated 250V ac, 0.5 Amps (alarm contacts) another 2 or 4 sets rated 250Vac, 2 Amps, 2 or 4 4-20mA signal circuits, probe and sensor connections all SELV (HART, fieldbus, profibus) – all part of component controller CM44x

Notes:

1. The above model is permanently connected, Pollution Degree 2, Measurement Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: 0 to 45C, 2000m max



Certificate: 70134179

Project: 70185140

Master Contract: 205557

Date Issued: 2018-07-16

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

ANSI/UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 70134179

Master Contract: 205557 (205557)

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
70185140	2018-07-16	Update report to cover adjustments to the model number and specifications of the power supply and revisions of the electric circuit plan.
70134179	2017-11-23	New certification for model CDC90, panel version.