



Certificate of Compliance

Certificate: 2742278

Master Contract: 205557

Project: 80144655

Date Issued: 2022-12-07

Issued To: Endress+Hauser Conducta GmbH & Co. KG
Dieselstraße 24
Gerlingen, Baden-Württemberg, 70839
Germany

Attention: Henryk Hofmann

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Andreas Kohler*
Andreas Kohler



PRODUCTS

CLASS 2252 06 - PROCESS CONTROL EQUIPMENT

CLASS 2252 86 - PROCESS CONTROL EQUIPMENT Certified to US Standards

Ultrasonic generator, model CYR52; rated 115V ac, 50W, 60 Hz or 230V ac, 50W, 50 Hz; Ambient rated -10 to +60 Deg. C. Enclosure rated CSA type 4X, IP 66/67.

Notes:

1. Rated: Pollution Degree 2, Overvoltage Category II.
2. Environmental Conditions: 2000 m max



Certificate: 2742278
Project: 80144655

Master Contract: 205557
Date Issued: 2022-12-07

APPLICABLE REQUIREMENTS

- | | |
|---|--|
| CAN/CSA-C22.2 No. 61010-1-12 | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements |
| UL Std. No. 61010-1 (3 rd Edition) | - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements |

Notes:

Products certified under Class C225206, C225286 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 2742278

Master Contract: 205557

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80144655	2022-12-07	Update to report 2742278 to cover minor changes of LOCC components without testing.
2742278	2014-07-14	New certification for ultrasonic generator, Model CYR52.