



# Certificate of Compliance

**Certificate:** 70134179 **Master Contract:** 205557 (205557)

**Project:** 70134179 **Date Issued:** 2017-11-23

**Issued to:** Endress+Hauser Conducta GmbH & Co. KG  
Dieselstrasse 24  
Gerlingen, 70839  
GERMANY  
Attention: Valery Minkousse

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



**Issued by:** *Dimcho Genov*  
Dimcho Genov

## **PRODUCTS**

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT-Certified to US Standards

Automated Sensor Cleaning and Calibration System, Liquiline Control CDC90

Rated 100 – 230 Vac, 50/60Hz, 50VA max;

One set of relay contacts rated 250V ac, 0.5 Amps (alarm contacts) another 2 or 4 sets rated 250Vac, 2 Amps, 2 or 4 4-20mA signal circuits, probe and sensor connections all SELV (HART, fieldbus, profibus) – all part of component controller CM44x

## **Notes:**

1. The above model is permanently connected, Pollution Degree 2, Measurement Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: 0 to 45C, 2000m max



**Certificate:** 70134179

**Project:** 70134179

**Master Contract:** 205557

**Date Issued:** 2017-11-23

---

**APPLICABLE REQUIREMENTS**

CSA Standards:

CAN/CSA-C22.2 No. 61010-1-12

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL Standards:

UL Std. No. 61010-1 (3<sup>rd</sup> Edition)

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



## *Supplement to Certificate of Compliance*

**Certificate:** 70134179

**Master Contract:** 205557 (205557)

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
70134179	2017-11-23	New certification for model CDC90, panel version.