

Company **Endress+Hauser SE+Co. KG, Hauptstraße 1, 79689 Maulburg**

declares as manufacturer, that the following product

Product **Cerabar**
PMP23-AA1N1HB3AJ+F3I7JAKB

is in conformity with the following provisions:

The roughness of the process isolating diaphragm (sensor membrane) is $R_a \leq 0.38 \mu\text{m}$ (15.00 μin).
Roughness of other wetted surfaces (excluded the weld seam) is $R_a \leq 0.76 \mu\text{m}$ (29.90 μin).

This declaration of compliance is valid for the listed products and accessories in delivery status.
The validity of this document for placing the product on the market expires 5 years after the date of issue.

Maulburg, 29-November-2018
Endress+Hauser SE+Co. KG



i.V. Marc Schlachter
Dept. Manager Pressure
Research & Development Devices