

Certificate of Compliance

Certificate: 2595034 Master Contract: 151079

Project: 80022663 **Date Issued:** 2020-04-22

Issued To: Endress+Hauser SE+Co. KG

Hauptstrasse 1

Maulburg, Baden-Württemberg, 79689

Germany

Attention: Arno Götz

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: Dimcho Genov Dimcho Genov



PRODUCTS

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT Certified to US Standards

Prosonic Model T FMU30- (followed by additional numbers and / or letters denoting specific features) is a permanently connected, non-contact Level Measuring Device. Depending on the sensor (1-1/2" or 2" Sensor), the measuring range is up to 8 m (26 ft) in fluids and up to 3,5 m (11 ft) in bulk solids.

Ultrasonic level measurement, stationary mounted, CSA Enclosure Type 6P / IP68.

An Electronic module FEU31 is used with ratings as follows: 14...35 VDC, 0.8W; 4...20 mA, two wire output; maximum ambient temperature 60 °C.



 Certificate: 2595034
 Master Contract: 151079

 Project: 80022663
 Date Issued: 2020-04-22

APPLICABLE REQUIREMENTS

CSA Standards:

CAN/CSA C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements - Third Edition

UL Standards:

ANSI/UL Std. No. 61010-1 (3rd Edition) - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



Supplement to Certificate of Compliance

Certificate: 2595034 Master Contract: 151079

The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.

Product Certification History

Project	Date	Description
80022663	2020-04-22	Update report 2595034 to cover evaluation to CSA C22.2 No. 61010-1-12 (3rd edition) and revise a drawing; under the CPC program.
0002595034	2013-04-04	Original Certification.