

CERTIFICATE OF CONFORMITY



1. **HAZARDOUS (CLASSIFIED) LOCATION ELECTRICAL EQUIPMENT PER US REQUIREMENTS**

2. **Certificate No:** FM20US0077X

3. **Equipment:** iTEMP type TMT180-abcdef
(Type Reference and Name) Head Transmitter

4. **Name of Listing Company:** Endress+Hauser Wetzer GmbH + Co. KG

5. **Address of Listing Company:** Obere Wank 1
D-87484 Nesselwang
Germany

6. The examination and test results are recorded in confidential report number:

3027190 dated 18th July 2006

7. FM Approvals LLC, certifies that the equipment described has been found to comply with the following Approval standards and other documents:

FM Class 3600:2018, FM Class 3611:2018, FM Class 3810:2018,
ANSI/UL 121201:2018, ANSI/ISA 61010-1:2012

8. If the sign 'X' is placed after the certificate number, it indicates that the equipment is subject to specific conditions of use specified in the schedule to this certificate.

9. This certificate relates to the design, examination and testing of the products specified herein. The FM Approvals surveillance audit program has further determined that the manufacturing processes and quality control procedures in place are satisfactory to manufacture the product as examined, tested and Approved.

10. Equipment Ratings:

Non-incendive for Class I, Division 2, Groups A, B, C and D, T*
Class I, Zone 2 IIC; T*
T* = T6 Ta = -40°C to +50°C, T5 Ta = -40°C to +65°C, T4 Ta = -40°C to +85°C

11. The marking of the equipment shall include:

Certificate issued by:

J.E. Marquedant
VP, Manager - Electrical Systems

13 July 2020

Date

To verify the availability of the Approved product, please refer to www.approvalguide.com

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmapprovals.com www.fmapprovals.com

SCHEDULE

US Certificate Of Conformity No: FM20US0077X



Member of the FM Global Group

Class I, Division 2, Groups A, B, C and D; T*, 14 13 00 111 NIFW

Class I, Zone 2 IIC, T* 14 13 00 111 NIFW

T* = T6 Ta = -40°C to +50°C, T5 Ta = -40°C to +65°C, T4 Ta = -40°C to +85°C

12. Description of Equipment:

General - The head transmitter iTEMP type TMT180 is a 2-wire transmitter with an analog output. It has measuring input for resistance thermometer Pt100 in 2-, 3- or 4-wire connection.

Ratings -

Non-Incendive Wiring Parameters:

Terminals	V _{max}	C _i	L _i
	(V)	(η F)	(mH)
1 and 2	30	144	0

iTEMP TMT180-abcdef. Head Transmitters

a = Approval: C or I.

b = Programming: 1 or 2.

c = Max. range, accuracy: not safety relevant

d = Configuration connection: 2, 3, or 4.

e = Configuration range: not safety relevant.

f = Model: A, B, or K.

13. Specific Conditions of Use:

- 1) The Temperature Transmitters shall be installed in a tool secured enclosure in compliance with the mounting, spacing and segregation requirements of the ultimate application.

14. Test and Assessment Procedure and Conditions:

This Certificate has been issued in accordance with FM Approvals US Certification Requirements.

15. Schedule Drawings

A copy of the technical documentation has been kept by FM Approvals.

16. Certificate History

Details of the supplements to this certificate are described below:

Date	Description
18 th July 2006	Original Issue.

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA

T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com

SCHEDULE

US Certificate Of Conformity No: FM20US0077X



Member of the FM Global Group

13 th July 2020	Supplement 1: Report Reference: – RR224071 dated 13 th July 2020. Description of the Change: Updated FM standards to the latest editions. Added ANSI/UL 121201 and ANSI/ISA 61010-1. Updated Certificate format with new unique certificate number. Added T6 Ta = -40°C to +50°C, T5 Ta = -40°C to +65°C,
----------------------------	---

FM Approvals

FM Approvals

FM Approvals

THIS CERTIFICATE MAY ONLY BE REPRODUCED IN ITS ENTIRETY AND WITHOUT CHANGE

FM Approvals LLC. 1151 Boston-Providence Turnpike, Norwood, MA 02062 USA
T: +1 (1) 781 762 4300 F: +1 (1) 781 762 9375 E-mail: information@fmaprovals.com www.fmaprovals.com