


Company **Endress+Hauser SE+Co. KG, Hauptstrasse 1, D-79689 Maulburg**
being the manufacturer, declares that the following devices


Device **Liquiphant**
FTL50, FTL51, FTL51B

may be used with a hydrogen content of 20 % in a natural gas storage plant and network infrastructure, at a maximum process temperature of 150 °C. No risk of technical issue or hydrogen diffusion is expected.

Maulburg, 20-Mai-2021
Endress+Hauser SE+Co. KG



i.V. Thorsten Springmann
Dept. Manager Level Continuous
Research & Development



i.V. Gerd Bechtel
Dept. Manager Level Limit
Research & Development