	M	OIML Certificate
	OIML Member State The Netherlands	Number R117/2007-A-NL1-21.01 revision 1 Project number 3147081 Page 1 of 6
	Issuing authority Person responsible:	NMi Certin B.V. M.Ph.D. Schmidt
Đ	Applicant and Manufacturer	Endress+Hauser Flowtec AG Kägenstrasse 7 4153 Reinach Switzerland
	Identification of the certified type	A measurement transducer Type: Promass F 300 DNxxx ^[1] ; Promass F 500 DNxxx ^[1] ; Promass O 300 DNxxx ^[1] ; Promass O 500 DNxxx ^[1] ; Promass X 300 DNxxx ^[1] ; Promass X 500 DNxxx ^[1] ; Promass Q 300 DNxxx ^[1] ; Promass Q 500 DNxxx ^[1] ;
	Characteristics	See page 2 and further.
	This OIML Certificate is	ssued under scheme A
	This Certificate attests to identified in the OIML Recommendation of th	e conformity of the above-identified type (represented by the sample(s) /pe Evaluation Report) with the requirements of the following International Organization of Legal Metrology (OIML): R 117-1:2007 "Dynamic measuring systems for liquids other than water"
	Accuracy class	0.3 / 0.5 / 1.0 / 1.5

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation identified above. This Certificate does not bestow any form of legal international approval.

This certificate and supporting reports comply with the requirements of OIML-CS-PD-07 clause 6.2.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Type Evaluation Report(s) is not permitted, although either may be reproduced in full.

[1] With xxx denoting the size of the Promass Q measurement sensor.

Issuing Authority

NMi Certin B.V.

The Netherlands

T +31 88 636 2332

Thijsseweg 11

2629 JA Delft

certin@nmi.nl

www.nmi.nl

NMi Certin B.V., OIML Issuing Authority NL1 11 April 2022

Certification Board

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

This document is digitally signed and sealed. The digital signature can be verified in the blue ribbon at the top of the electronic version of this certificate.





(+)



OIML Member State

The Netherlands

OIML Certificate

Number R117/2007-A-NL1-21.01 revision 1 Project number 3147081 Page 2 of 6

The conformity was established by the results of tests and examinations provided in the associated reports:

Report number	Issue date	Number of pages						
	Measurement sensor							
PF/6491	2 July 1996	77						
CVN-302404-01 rev. 1	27 June 2003	15						
CPC-407631-1	31 March 2005	42						
CPC-412432-1	31 March 2005	32						
CPC-10200667-1	9 August 2010	7						
NMi-10201056-1	29 March 2012	7						
NMi-11200539-01	2 April 2012	11						
NMi-12200149-1	24 May 2012	11						
NMi-12200688-01	3 December 2013	30						
NMi-14200053-01	16 April 2014	18						
NMi-15200446-01	22 June 2016	11						
NMi-15200323-01a	12 February 2021	94						
NMi-1901704-01	31 August 2018	17						
NMi-1902055-01 👘	31 August 2018	26						
NMi-2389303-01	12 May 2020	31						
NMi-2389303-02	12 February 2021	17						
NMi-2389303-03	12 February 2021	12						
NMi-2463103-01	12 February 2021	17						
Promass 300/500 electronics								
NMi-16200475-01	22 December 2016	247						
NMi-1901185-01	1 November 2017	269						
NMi-2202829-01	3 December 2019	261						



OIML Certificate

OIML Member State The Netherlands



Number R117/2007-A-NL1-21.01 revision 1 Project number 3147081 Page 3 of 6

Characteristics of the measurement transducer

In Tables 1 to 6, the general characteristics of the measuring instrument are presented. The construction of the measurement transducer is recorded in documentation folders TC7149-7 for the measurement sensor and TC10822-3 for the electronics.

Table 1 General characteristics applicable to all Promass measurement sensors

- Density range:
- Maximum viscosity:
- Accuracy class:
- Environmental classes:
- 400 ... 1400 kg/m³ 1000 mPa·s 0.3; 0.5; 1.0 and 1.5 M3 / E2 / H3 -40 ... +55 °C
- Ambient temperature range:Intended for the measurement of:
 - Oil and oil products, chemicals, potable liquids, liquefied gases under pressure measured at temperatures above -10 °C, liquified gases under pressure measured below -10 °C, liquefied carbon dioxide and liquefied natural gas (LNG).

Table 2 Specific characteristics of the Promass F measurement sensors

Sensor size	DN8	DN15	DN25	DN40	DN50	DN80
Maximum flow rate [kg/min]	30	100	300	700	1000	3000
Minimum flow rate [kg/min] ^[2]	1,5	5	15	37,5	58,3	150
Minimum flow rate [kg/min] ^[3]	0,75	2,5	7,5	18,75	29,15	75
Minimum Measured Quantity [kg]	2	5	20	20	20	200

Sensor size	DN100	DN150	DN250		
Maximum flow rate [t/h]	270	720	2200		
Minimum flow rate [t/h] ^[2]	14	32	90		
Minimum flow rate [t/h] ^[3]	7	16	45		
Minimum Measured Quantity [kg]	200	500	1000		

Further characteristics of the Promass F:

Accuracy Class	0.3	0.5	1.0	1.5	
Maximum pressure					
Temperature range liquid for mass measurement	-	-10 °C +200 °C			
Temperature range liquid for density and volume measurement		-10 °C +85 °C			

[2] For accuracy class 0.3 and 0.5

[3] For accuracy class 1.0 and 1.5

 \bullet



OIML Member State The Netherlands



Number R117/2007-A-NL1-21.01 revision 1 Project number 3147081 Page 4 of 6

Table 3 Specific characteristics of the Promass X measurement sensors

Sensor size	DN350			
Maximum flow rate [t/h]	3353			
Minimum flow rate [t/h] ^[4]	137			
Minimum flow rate [t/h] ^[5]	68,5			
Minimum Measured Quantity [kg]	1000			

Further characteristics of the Promass X:

Accuracy Class	0.3	0.5	1.0	1.5	
Maximum pressure					
Temperature range liquid for mass measurement	-10 °C +180 °C			-10 °C +85 °C	
Temperature range liquid for density and volume measurement		-10 °C +85 °C			

Table 4 Specific characteristics of the Promass O measurement sensors

Sensor size	DN8	DN15	DN25	DN40	DN50	DN80
Maximum flow rate [kg/min]	30	100	300	700	1000	3000
Minimum flow rate [kg/min]	1,5	5	15	37,5	58,3	150
Minimum Measured Quantity [kg]	2	5	20	20	20	200

Sensor size	DN100	DN150	DN250		
Maximum flow rate [t/h]	270	720	2200		
Minimum flow rate [t/h]	14	32	90		
Minimum Measured Quantity [kg]	200	500	1000		

Further characteristics of the Promass O:

Accuracy Class	0.3	0.5	1.0	1.5
Maximum pressure				
Temperature range liquid for mass measurement	-10 °C	+200 °C	NA	NA
Temperature range liquid for density and volume measurement	-10 °C +85 °C		NA	NA

[4] For accuracy class 0.3 and 0.5

[5] For accuracy class 1.0 and 1.5



OIML Member State The Netherlands



Number R117/2007-A-NL1-21.01 revision 1 Project number 3147081 Page 5 of 6

Table 5 Specific characteristics of the Promass Q measurement sensors

Sensor size	DN25	DN50	DN80	DN100	
Maximum flow rate [t/h]	20	80	200	400	
Minimum flow rate [t/h] ^[6]	0,45	2	6	14	
Minimum flow rate [t/h] ^[7]	0,225	1	3	7	
Minimum Measured Quantity [kg]	10	20	100	200	

Further characteristics of the Promass Q:

Accuracy Class	0.3	0.5	1.0	1.5
Maximum pressure	100 bar(g)			
Temperature range liquid for mass measurement	-10 °C +200 °C			-200 °C +90 °C
Temperature range liquid for density and volume measurement	-	NA		

•



[6] For accuracy class 0.3 and 0.5[7] For accuracy class 1.0 and 1.5



OIML Member State The Netherlands



Number R117/2007-A-NL1-21.01 revision 1 Project number 3147081 Page 6 of 6

Table 6 General characteristics of the Promass 300 and Promass 500 electronics

Environmental classes		M3 / E2 / H3							
	Ambient temperature range			-40+55 °C; condensing humidity					
Power supply voltage			24 VDC 100240 VAC, 5060 Hz 24 VDC / 100240 VAC, 5060 Hz						
S	oftware identific	ation							
	Version	Chec	ecksum			Version	Checksum		
	number	Modbus	Har	t		number	Modb	us	Hart
	01.00.02 ^[8]	0xE87F	0x321F			01.02.00	0x564	15	-
	01.00.03 ^[8]	0x79B5	0x1585			01.02.01	0x5645		-
	01.00.04	0xE109	0xB0	75		01.02.02	0x0A9	92	-
	01.01.01	0xA476	0x977	7D		01.02.03	0xECE	3	-
	01.01.02	0x2AAB	0xED4	44		01.05.00	0xA9I	E	0xB4A1
	01.01.03	0x6A37	0x86I	FC		01.05.01	0x2B9	95	0x59D4
	01.01.04	0x6D79	0x67	4		01.05.02	0xF1E	37	0xE6B5
	01.01.05	0x4670	0x559	9B		01.05.03	-		0xD79D
	01.01.06	-	0x089	91					
	01.01.07	-	0xB7I	B2					

The Promass 300 and Promass 500 flow transmitter is solely to be used in combination with dynamic mass meters (Coriolis meters) of brand Endress+Hauser.

Certificate history

This revision replaces the previous version.

Revision	Date	Description of the modification		
Initial	18 March 2021	-		
1	11 April 2022	Additional software versions for the Promass 300/500 flow transmitter added.		

