

## JP Food Contact Material - Positive List Declaration of Compliance

**Company** Endress+Hauser Conducta GmbH+Co. KG  
 Dieselstraße 24, 70839 Gerlingen, Germany

as a manufacturer of products for liquid analysis declares, that the product

**Product** OUA260-\*A1\*\*\*x1\*\*y  
 x = B, C, D  
 y = A, B

**Regulations** is conform to the Positive List in appendix I of Japan´s MHLW Notification No. 370 (Notification No. 196 of 2020) and its latest reorganization published on December 24, 2021.

The following process contact parts are used:

Process contact material	Extended order code	Material group	Material
O-ring	OUA260-*A1***x1**y x=B, C, D y=A, B	Rubber	EPDM
Flow cell	OUA260-*A1***x1**y x=B, C, D y=A, B	Metal	Stainless steel 316L / 1.4435
Window	OUA260-*A1***B1**y y=A, B	Glass	Quartz
Window	OUA260-*A1***C1**y y=A, B	Other material	Sapphire (99,997 % Al <sub>2</sub> O <sub>3</sub> )
Window	OUA260-*A1***D1**y y=A, B	Glass	Borosilicate

**Further information** The product mentioned above does not contain any plastics that fall under the positive list regulations of the Japanese MHLW Notification No. 370.

## JP Food Contact Material - Positive List Declaration of Compliance

Remarks            The product has to be cleaned before use.

This declaration relates exclusively to the positive list. No further tests were carried out.

The validity of this document expires on June 1, 2025.

This declaration of conformity is to emphasize that the customer is obliged to test the product with regard to its suitability in the application. This declaration of conformity is only valid for the listed product as supplied to customer.


Endress+Hauser Conducta GmbH+Co. KG  
Dieselstraße 24, 70839 Gerlingen, Germany

22.07.2024



i. V. Uwe Rößiger

Head of Group Certificates Approv. Tech.  
G. Certificates & Approvals



i. V. Dr. Julia Groß

Senior Specialist  
Tech. W. R. Sensors Oxygen