

2. SUPPLEMENT

according to Directive 94/9/EC Annex III.6

to EC-TYPE-EXAMINATION CERTIFICATE PTB 02 ATEX 2062 X

(Translation)

Equipment: Pressure Transducer Cerabar T, types PMP 131/135 and
Measuring Transducer P40/41

Marking:  II 1/2 G Ex ib IIC T6 Ga/Gb bzw. II 2 G Ex ib IIC T6 Gb

Manufacturer: Endress + Hauser GmbH+Co. KG

Address: Hauptstraße 1, 79689 Maulburg, Germany

Description of supplements and modifications

In the future the Pressure Transducer Cerabar T types PMP 131/135 and Measuring Transducer P40/41 may also be manufactured and operated according to the test documents listed in clause 3 of the test report.

The modification effects the Pressure Transducer Cerabar T types PMP 131/135 and concerns the update of the applied standards.

The marking and all further specifications of the EC-Type-Examination Certificate as well as the 1st supplement apply without changes for this 2nd supplement.

Special conditions for safe use

All special conditions for safe use of the EC-Type-Examination Certificate as well as the 1st supplement apply without changes for this 2nd supplement.


Applied standards

EN 60079-0:2012, EN 60079-11:2012, EN 60079-26:2007

Test report: PTB Ex 14-24002

Zertifizierungssektor Explosionsschutz
On behalf of PTB:

Braunschweig, March 3, 2014


Dr.-Ing. U. Johannsmeyer
Direktor und Professor

