



Certificate

PROFIBUS Nutzerorganisation e.V. grants to

Endress+Hauser GmbH + Co. KG
Hauptstraße 1; 79689 Maulburg; Germany

the Certificate No.: **Z00628** for the PROFIBUS Slave:

Product Name: Cerabar M
Revision: Profile 3.0; SW/FW: 1.2; HW: 3.0
GSD: EH3X151C.gsd; File version: 0.3
PA139700.gsd

This certificate confirms that the product has successfully passed the certification tests with the following scope:

<input checked="" type="checkbox"/>	DP-V0	MS0, Freeze; Set_Slave_Add
<input checked="" type="checkbox"/>	DP-V1	MS2
<input checked="" type="checkbox"/>	Profile	PA Devices V3.0
<input checked="" type="checkbox"/>	Physical Layer	MBP

Test Report Number: itm 628 PA 02/01
Authorized Test Laboratory: itm, München, Germany


The tests were executed in accordance with the following documents:

"Test Specifications for PROFIBUS DP Slaves, Version 3.0 from November 2005" and
"Test Specification for PA Devices, Profile 3.0" Version 4.1 from November 2004".

This certificate is granted according to the document "Framework for testing and certification of PROFIBUS products".

For all products that are placed in circulation by December 11, 2011 the certificate is valid for life.




.....
(Official in Charge)

Board of PROFIBUS Nutzerorganisation e. V.



(J. Freitag)



(Prof. K. Bender)