



# Certificate of Compliance

Certificate: 1578782

Master Contract: 200600

Project: 1578782

Date Issued: 2004/11/09

Issued to: Endress + Hauser Wetzer Gmbh Co. KG

Obere Wank 1  
Nesselwang, 87484  
Germany  
Attention: Mr. Michael Pfanzelt

*The products listed below are eligible to bear the CSA Mark shown*



Issued by:

A handwritten signature in black ink, appearing to read 'E. Foo'.

E. Foo, C.E.T.

Authorized by: Nick Alfano, Operations  
Manager

A handwritten signature in black ink, appearing to read 'N. Alfano'.

## PRODUCTS

**CLASS 2252 03** - PROCESS CONTROL EQUIPMENT

- Compact Thermometer, Type TSM 470, TSM 470F, TSM 470G and TSM 470P Series, SELV or Class 2 supply rated 10 to 35Vdc, 23mA. Ambient temperature = -40°C to +85°C.

## APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 1010.1-92 - Safety requirements for electrical Amends 1 and 2 equipment for measurement, control and laboratory use.



CSA INTERNATIONAL

## *Supplement to Certificate of Compliance*

**Certificate:** 1578782

**Master Contract:** 200600

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
1578782	2004/11/09	CSA Certification of Compact RTD Transmitter to C22.2 No. 1010.1-92 for Installation in Ordinary Locations