



CSA INTERNATIONAL

# Certificate of Compliance

Certificate: 1681249

Master Contract: 200600

Project: 1681249

Date Issued: 2005/11/25

Issued to: Endress + Hauser Wetzler Gmbh Co. KG

Obere Wank 1  
Nesselwang, 87484  
Germany  
Attention: Michael Pfanzelt

*The products listed below are eligible to bear the CSA Mark shown*



Issued by: Sam Zaffino

Authorized by: Patricia Pasemko, Operations  
Manager

## PRODUCTS

**CLASS 2252 05 - PROCESS CONTROL EQUIPMENT**

Controller, Model: Memo Graph S, Rack or panel mounted equipment, Rated 115-230V, 50/60Hz, 25VA input; Relay contacts rated 230 Vac, 3.0 Amps; signal circuits all SELV, Maximum ambient 50 Deg. C; Pollution degree 2, Installation Category II.

## APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements



CSA INTERNATIONAL

## *Supplement to Certificate of Compliance*

Certificate: 1681249

Master Contract: 200600

*The products listed, including the latest revision described below, are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
1681249	2005/11/25	Certification of Model Memo Graph S to C22.2 No. 61010-1-04