



# Certificate of Compliance

**Certificate:** 1593317

**Master Contract:** 151079

**Project:** 70122700

**Date Issued:** 2017-03-24

**Issued to:** Endress + Hauser GmbH + Co. KG  
Hauptstrasse 1  
Maulburg, Baden-Württemberg 79690  
GERMANY  
Attention: Arno Götz

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.*



**Issued by:** Sam Zaffino  
Sam Zaffino

## **PRODUCTS**

CLASS - C225206 - PROCESS CONTROL EQUIPMENT

CLASS - C225286 - PROCESS CONTROL EQUIPMENT-Certified to US Standards

Level Limit Switch, 'Soliphant T' models FTM20 and FTM21, with module FEM22 rated 10-45Vdc, 350 mA max. and module FEM24 rated 19-253Vac, 50/60 Hz or 19-55Vdc, 1.3 VA max.

## **APPLICABLE REQUIREMENTS**

CAN/CSA-C22.2 No. 61010-1-12 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

ANSI/UL Std. No 61010-1 (3rd Edition)- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements



## *Supplement to Certificate of Compliance*

**Certificate:** 1593317

**Master Contract:** 151079

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
70122700	2017-03-24	Update to 1593317 to cover evaluation to CAN/CSA 61010-1-12, corrections and alt construction
1593317	2004-11-05	Original evaluation