



Manufacturer Information

for users regarding software updates
 (following the NAMUR recommendation 53)

1. Type of device

- Field device / signal processing device
- Monitoring- / operation system / hand held terminal etc.
- Modem / interface

Manufacturer : Endress+Hauser Flowtec AG
 Device : Coriolis Flow Measuring System
 Type : Promass 200 HART (0x54)
 Order code : 8E2B**_*A*****
 8E2B**_*B*****
 8E2B**_*C*****

2. Software

Previous firmware version : 1.01.03
 New firmware version : 1.02.00
 How can the previous firmware version be identified? :

- Local display
- Nameplate for connections
- Integrated Webserver
- Configuration and service software e.g. FieldCare

 Description of the modifications in comparison to the previous firmware version :

- Pulse output added

3. Compatibility

Is the operating tool compatible with the installed device firmware?

- Yes
- No, description:

An update of the DD's (device description) is required for all operating tools in use.



Manufacturer Information

for users regarding software updates
(following the NAMUR recommendation 53)

Is a firmware update generally recommended?

- Yes:
 No, reason:

If the additional properties are not required for the operation of the measurement system, then no firmware update is necessary.

4. Instruction manual

Is a new instruction manual necessary due to the modification of the firmware?

- Yes
 No

Which manual corresponds to the new firmware?

Device	Communication options	Manual	Identification
Promass 200	HART	Operating instructions	BA01027D/06/en/15.14

The new instruction manuals can be referred in Internet:

- <http://www.endress.com> - area „DOWNLOAD“
- declaration of the device and kind of manual

5. Price

Change in price of device in comparison with the predecessor version?

- Yes, new list price and update costs (without installation) are enclosed
 No